



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

| | | |
|---|--|----------------------------------|
| Chemical Test Operator Name: <u>Chad L. Harmon</u> | | |
| Location: <u>FYPD</u> | | |
| Intoxilyzer® Serial Number: <u>80-004954</u> | | |
| Gas Lot Number: <u>709352</u> | Gas Cylinder Number: <u>26</u> | Expiration Date: <u>12-23-11</u> |
| <input type="checkbox"/> New Cylinder Installed | <input checked="" type="checkbox"/> Cylinder Not Changed | |
| Instrument Test Results (Report to 3 Digits; Example, 0.081%) | | |
| Test 1: 0. <u>081</u> % | | |
| Test 2: 0. <u>080</u> % | | |
| Test 3: 0. <u>081</u> % | | |
| (ATTACH TEST RECORD) | | |

[Signature]
Chemical Test Operator Signature

12-15-10
Date

Debi Kashur
Reviewed By

17 Dec 2010
Date

80-004954
P. 63 of 44
[Signature]


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = FYPD 8164.13.00 06/09
12/15/2010 20:20

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 20:21 |
| 02 Std. Gas | 0.081 | 20:21 |
| 03 Room Air | 0.000 | 20:22 |
| 04 Std. Gas | 0.080 | 20:22 |
| 05 Room Air | 0.000 | 20:23 |
| 06 Std. Gas | 0.081 | 20:23 |
| 07 Room Air | 0.000 | 20:24 |

Lot No = 709352
Cyl No = 26
Exp Date = 12/23/2011
County = 61 Oper No. = 010606



Operator Signature
CHAD L. HARMON

Remarks:

Form 106-I8000

80-004954
P. 64 of 64
sam