

**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**

Serial No.: 80-002669 Instrument Location: Cavalier, ND

Reason for Install/Repair: \_\_\_\_\_ Install After Lab Repair/Inspection  
 Other (Specify) \_\_\_\_\_

Check When Done:

- A. Surge Protector Installed/Properly Grounded.
- B. Telephone Line Connected to Intoxilyzer®.
- C. Breath Tube Heated.
- D. Review/Enter Preliminary Data Entry (Level 2, Function E).
- E. Review/Enter Gas Setup (Level 1, Function S).
- F. Print Test (Level 1, Function P). (Attach Test Record.)
- G. ACA Test (Level 1, Function C). (Attach Test Record.)
- H. RFI Test (Use CMS Mode-First Room Air). (Attach Test Record.)
- I. Repair and/or Maintenance Performed: \_\_\_\_\_

12/16/09  
Date

*P. M. Kelly*  
Field Inspector's Signature

24 Dec 2009  
Date

*Debi Keshun*  
Reviewed By

80-002669  
P. 1 of 4

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-002669  
Location = CAVA                    8164.13.00 06/09  
12/16/2009                            22:31

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            659358  
Standard Cyl #?            34  
Standard Expiration?       09/03/2011  
Oper No?                    120547

Flow Cal. Date:            09/10/2007  
Slope                       702  
Intercept                  -608746

IR Calibration Date:       12/11/2009  
                                    3um                    9um

0th Coef(*100):	-25342	-12745
1st Coef(*100):	257733	131681
2nd Coef(*100):	2435	984
H2O adj(mg/l*10k):	603	411

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
PHIL WOLF

Remarks:

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-002669  
Location = CAVA      8164.13.00 06/09  
12/16/2009      22:32

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	22:32
02 Std. Gas	0.077	22:33
03 Room Air	0.000	22:33
04 Std. Gas	0.077	22:34
05 Room Air	0.000	22:34
06 Std. Gas	0.078	22:34
07 Room Air	0.000	22:35


Lot No = 659358

Cyl No = 34

Exp Date = 09/03/2011

County = 34

Oper No. = 120547

  
Operator Signature  
PHIL WOLF

Remarks:

Form 106-I8000

80-002669  
P. 3 of 4 JAH


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-002669  
Location = CAVA      8164.13.00 06/09  
12/16/2009      22:36

Test	AC	Time
01 Room Air	RFI*	22:37
02 Room Air	RFI*	22:37

\*Invalid Test  
Inhibited - RFI

Sub Name = SMITH, JOHN T  
Sub DOB = 12/16/1976  
Sub Sex = Male      Weight = 176  
Test = DUI      Cit = 1111111  
Dr. Lic. = ND/SMI761234  
Lot No = 659358  
Cyl No = 34  
Expiration Date = 09/03/2011  
County = 34      Oper No. = 120547

  
\_\_\_\_\_  
Operator Signature  
PHIL WOLF

Remarks:

Form 106-I8000

80-002669  
P. 4 of 4 12/16/09