

ETHANOL BREATH STANDARD CYLINDER REPORT

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Phil Wolf

Location: Luxatier, MD

Intoxilyzer Serial No.: 80-002669

Lot Number: 659357

Cylinder Number: 9

Expiration Date: 9/3/11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.077 %

Test 2: 0.077 %

Test 3: 0.076 %

(ATTACH TEST RECORD)

P. Wolf
Chemical Test Operator Signature

12/14/09
Date

Deb Keshur
Reviewed By

24 Dec 2009
Date

80-002669
P. Wolf
AK


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = CAVA 8164.13.00 06/09
12/16/2009 22:39

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	22:40
02 Std. Gas	0.077	22:40
03 Room Air	0.000	22:40
04 Std. Gas	0.077	22:41
05 Room Air	0.000	22:41
06 Std. Gas	0.078	22:42
07 Room Air	0.000	22:42

Lot No = 659358
Cyl No = 34
Exp Date = 09/03/2011
County = 34 Oper No. = 120547



Operator Signature
PHIL WOLF

Remarks:

Form 106-18000

80-002669
R 2 of 2
JPM