

03/11/2010 14:20 NDOAG - Crime Laboratory Div

(FAX)701 328 6185

P.002/006

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program

Form: BrF-105-G

ETHANOL BREATH STANDARD CYLINDER REPORT

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Mark W. McDonald

Location: Valley City PD

Intoxilyzer Serial No.: 80-004184

Lot Number: 659358

Cylinder Number: 47

Expiration Date: 09/03/2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.081 %

Test 2: 0.082 %

Test 3: 0.081 %

(ATTACH TEST RECORD)

[Signature]
Chemical Test Operator Signature

12/17/09
Date

Deb Kashur
Reviewed By

12 Mar 2010
Date

80-004184
73/24 FR


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = VALL 8164.13.00 06/09
12/17/2009 12:45

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:46
02 Std. Gas	0.081	12:46
03 Room Air	0.000	12:47
04 Std. Gas	0.082	12:47
05 Room Air	0.000	12:48
06 Std. Gas	0.081	12:48
07 Room Air	0.000	12:49

Lot No = 659358
Cyl No = 47
Exp Date = 09/03/2011
County = 02 Oper No. = 119502



Operator Signature
MARK W. MCDONALD

Remarks:

Form 106-I8000

80-004184
7/24 FM