

INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

Serial No.: 80-004188 Instrument Location: Cavalier County SO - hangdown

Reason for Install/Repair: Install After Lab Repair/Inspection

Other (Specify) _____

Check When Done:

- A. Surge Protector Installed/Properly Grounded.
- B. Telephone Line Connected to Intoxilyzer®.
- C. Breath Tube Heated.
- D. Review/Enter Preliminary Data Entry (Level 2, Function E).
- E. Review/Enter Gas Setup (Level 1, Function S).
- F. Print Test (Level 1, Function P). (Attach Test Record.)
- G. ACA Test (Level 1, Function C). (Attach Test Record.)
- H. RFI Test (Use CMS Mode-First Room Air). (Attach Test Record.)
- I. Repair and/or Maintenance Performed: _____

12/17/2009
Date

David J. Zew
Field Inspector's Signature

24 Dec 2009
Date

Deb Kashner
Reviewed By

80-004188
P.14/4 JEM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004188
Location = LANG 8164.13.00 06/09
12/17/2009 11:15

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 659358
Standard Cyl #? 53
Standard Expiration? 09/03/2011
Oper No? 067518

Flow Cal. Date: 06/04/2009
Slope 689
Intercept -630047

IR Calibration Date: 12/11/2009
 3um 9um

0th Coef(*100):	-36132	-37983
1st Coef(*100):	283422	138305
2nd Coef(*100):	-135	610
H2O adj(mg/l*10k):	688	672

***** Printer Test End *****



Operator Signature
DAVID J. ZEIS

Remarks:


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004188
Location = LANG 8164.13.00 06/09
12/17/2009 11:18

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:20
02 Std. Gas	0.080	11:20
03 Room Air	0.000	11:21
04 Std. Gas	0.079	11:21
05 Room Air	0.000	11:22
06 Std. Gas	0.079	11:22
07 Room Air	0.000	11:22

Lot No = 659358
Cyl No = 53
Exp Date = 09/03/2011
County = 10 Oper No. = 067518



Operator Signature
DAVID J. ZELS

Remarks:

Form 106-I8000

80-004188
P. 3 of 4 FM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004188
Location = LANG 8164.13.00 06/09
12/17/2009 11:27

Test	AC	Time
01 Diagnostic	OK	11:30
02 Room Air	0.000	11:30
03 *Subject Test	RFI*	11:31
04 Room Air	0.000	11:31

*Invalid Test
Inhibited - RFI

Sub Name = TEST, TEST T
Sub DOB = 01/01/2009
Sub Sex = Male Weight = 150
Test = OTH Cit = 1
Dr. Lic. = ND/111111111
Lot No = 659358
Cyl No = 53
Expiration Date = 09/03/2011
County = 10 Oper No. = 067518

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DAVID J. ZEIS

Remarks:

Form 106-I8000

80-00418P
P. 4 of 4 SEM