

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program

Form: BrF-105-G

ETHANOL BREATH STANDARD CYLINDER REPORT

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: David J. Zeis

Location: Cavalier County SO Langdon

Intoxilyzer Serial No.: 80-004188

Lot Number: 659358

Cylinder Number: 53

Expiration Date: 09/03/11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 079 %

Test 2: 0. 078 %

Test 3: 0. 079 %

(ATTACH TEST RECORD)

David J. Zeis

Chemical Test Operator Signature

12/17/09

Date

Deb Kashner

Reviewed By

24 Dec 2009

Date


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004188
Location = LANG 8164.13.00 06/09
12/17/2009 11:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:36
02 Std. Gas	0.079	11:37
03 Room Air	0.000	11:37
04 Std. Gas	0.078	11:37
05 Room Air	0.000	11:38
06 Std. Gas	0.079	11:38
07 Room Air	0.000	11:39

Lot No = 659358
Cyl No = 53
Exp Date = 09/03/2011
County = 10 Oper No. = 067518



Operator Signature
DAVID J. ZEIS

Remarks:

Form 106-I8000

80-004188
p. 2 of 2 Jm