



## Intoxilyzer 8000 Initial Inspection

<b>Serial Number</b>	<b>80-004201</b>
----------------------	------------------

<input checked="" type="checkbox"/> A. Power On	<input checked="" type="checkbox"/> 5 min. countdown	<input checked="" type="checkbox"/> 1 min.	
<input checked="" type="checkbox"/> B. Security Checks.	<input checked="" type="checkbox"/> Level 1	<input checked="" type="checkbox"/> Level 2 <input checked="" type="checkbox"/> Level 3	
<input checked="" type="checkbox"/> C. Gas Cylinder Setup			
<input checked="" type="checkbox"/> D. Printer. Level 1 Func P	<input checked="" type="checkbox"/> Paper Feed	<input checked="" type="checkbox"/> Internal <input checked="" type="checkbox"/> External	
<input checked="" type="checkbox"/> E. Display is working.			
<input checked="" type="checkbox"/> F. Configure Start Test Switch. Level 3 Func S Sub T			
<input checked="" type="checkbox"/> G. Setup Preliminary Data. Level 2 Func E			
<input checked="" type="checkbox"/> H. Configure General Setup. Level 3 Func S Sub G.			
<input checked="" type="checkbox"/> I. Configure Port. Level 3 Func S Sub N			
<input checked="" type="checkbox"/> J. Check Camera. ABA Mode	<input checked="" type="checkbox"/> Oper ID	<input checked="" type="checkbox"/> ND Lic. <input checked="" type="checkbox"/> Hot Key "R"	
<input checked="" type="checkbox"/> K. Diagnostic Test. Level 2 Func D			
<input checked="" type="checkbox"/> L. Check Ser Num / Version Level 2 Func V	80-004201 Serial No. <u>8164.13</u> Version <u>1362.14</u> Display <u>7</u> CPLD		
<input checked="" type="checkbox"/> M. Detector Voltages. Level 3 Func D Sub D	3 <u>12800</u> 9 <u>12823</u>	s <u>1</u> s <u>3</u>	C <u>47.66</u> B <u>45</u> <u>0</u> RFI Ind.
<input checked="" type="checkbox"/> N. Hot Keys	<input checked="" type="checkbox"/> F5 Tone <input checked="" type="checkbox"/> F6 Green LED <input checked="" type="checkbox"/> F7 Red LED <input checked="" type="checkbox"/> F9 Fan <input checked="" type="checkbox"/> F10 Pump <input checked="" type="checkbox"/> F11 Sol Output		
<input checked="" type="checkbox"/> O. RFI Dectector			
<input checked="" type="checkbox"/> P. Atmospheric Monitor. Level 3 Func D Sub A.	P <u>1018</u>	CF <u>0.99</u>	

08/09

80-004201  
 P. 1/2/21  
 FM

OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION  
 Toxicology Section/Breath Alcohol Program  
 Instrument Certification and Maintenance Unit

BrF-005-8

Q. Flow Monitor. Level 3 Func D Sub F. R 186 F 0.000 L/s

R. Temperature Monitor. Level 3 Func D Sub T  
 C 47.0 A 49.0 I 0.0  
 B 45.1 V 0.0

S. Gas Tank Monitor. Level 3 Func D Sub G  
 T. Configure Wet Bath Sim. Level 1 Func S  
168 psi Intox 170 psi Reg

U. Zero Alcohol Test Level 1 Func C  
DR3409 Sim SN      Lot 0.000 Rep Value  
0.000 AC 0.000 AC 0.000 AC

V. Low AC Test. Level 1 Func C  
DR5144 Sim SN      Lot 0.019 Rep Value  
0.019 AC 0.019 AC 0.019 AC

W. Linearity Test. Level 1 Func C  
DR5187 Sim SN      Lot 0.285 Rep Value  
0.285 AC 0.286 AC 0.285 AC

X. Interferent Test. Level 1 Func B  
 Acetone DR3845 Sim SN      Lot 0.05% Conc  
 Display  Printout  
 Methanol 610573 Sim SN      Lot 0.10% Conc  
 Display  Printout  
 IPA DR5191 Sim SN      Lot 0.10% Conc  
 Display  Printout

Y. Ambient Test  
1122 Display 1122 Printout

Z. Invalid Sample Test  
 Display  Printout

AA. RFI Detector Test. CMS Mode  
 Display  Printout

BB. Configure Dry Gas Standard. Level 1 Func S

CC. Stability Test at 0.080AC. Level 3 Func T Sub S  
0.0796 Ave. 0.0006 Std Dev 0.76% Rel Std Dev

DD. Maintenance/Repairs Software upgrade 8/6/4.13.

22 Dec 2009  
 Date  
30 Dec 2009  
 Date

D. Kashur  
 Signature  
Mary Beaman  
 Reviewed by

TOXL  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-004201  
~~12/23/2009~~ 13:01:06  
 12/22/2009 DK  
 Auto Calibration

	<<<<< 3um >>>>>		<<<<< 9um >>>>>	
-----				
Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1				
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	0.1400	(0.0000)	0.2290	(0.0040)
Sample #2	0.0980	(0.0500)	0.2380	(0.0030)
Sample #3	0.1010	(0.0700)	0.2270	(0.0240)
Sample #4	0.1320	(0.0700)	0.2390	(0.0040)
Avg % Abs	0.1103	(0.0633)	0.2347	(0.0103)
STD DEV	0.0188	(0.0115)	0.0067	(0.0118)
REL STD DEV	17.061	(18.232)	2.837	(114.641)
-----				
Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1				
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	0.7730	(0.0000)	1.6310	(-0.0090)
Sample #2	0.7990	(0.0000)	1.6320	(-0.0130)
Sample #3	0.7810	(0.0230)	1.6290	(-0.0070)
Sample #4	0.7810	(0.0410)	1.6310	(0.0140)
Avg % Abs	0.7870	(0.0213)	1.6307	(-0.0020)
STD DEV	0.0104	(0.0206)	0.0015	(0.0142)
REL STD DEV	1.320	(96.332)	0.094	(708.872)
-----				
Solution = 0.107 g/210L or 0.5095 mg/l, Samples = 4, Discarded = 1				
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	1.9460	(0.0000)	4.0080	(0.0000)
Sample #2	1.9850	(0.0000)	4.0480	(-0.0010)
Sample #3	1.9550	(0.0140)	3.9890	(0.0280)
Sample #4	1.9460	(0.0230)	4.0140	(0.0250)
Avg % Abs	1.9620	(0.0123)	4.0170	(0.0173)
STD DEV	0.0204	(0.0116)	0.0296	(0.0159)
REL STD DEV	1.041	(93.975)	0.737	(92.007)
-----				
Solution = 0.147 g/210L or 0.7000 mg/l, Samples = 4, Discarded = 1				
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	2.5680	(-0.0120)	5.3100	(-0.0230)
Sample #2	2.5950	(-0.0080)	5.2810	(0.0080)
Sample #3	2.6200	(0.0070)	5.3340	(0.0000)
Sample #4	2.6090	(-0.0060)	5.3320	(0.0150)
Avg % Abs	2.6080	(-0.0023)	5.3157	(0.0077)
STD DEV	0.0125	(0.0081)	0.0300	(0.0075)
REL STD DEV	0.480	(349.051)	0.565	(97.899)
-----				
Solution = 0.285 g/210L or 1.3571 mg/l, Samples = 4, Discarded = 1				
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	5.0320	(0.0090)	10.0180	(0.0100)
Sample #2	5.0220	(0.0270)	10.0010	(0.0510)
Sample #3	5.0230	(0.0390)	10.0280	(0.0470)
Sample #4	5.0270	(0.0380)	10.0290	(0.0460)
Avg % Abs	5.0240	(0.0347)	10.0193	(0.0480)
STD DEV	0.0026	(0.0067)	0.0159	(0.0026)
REL STD DEV	0.053	(19.207)	0.159	(5.512)
-----				

*Calibration  
 Deb Kashner  
 p. 1 of 2*

*80-004201  
 p. 3 of 21, 1/24*

TOXL  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-004201  
~~12/23/2009~~ 13:01:06  
 12/22/2009 DK  
 Auto Calibration

<<<<< 3um >>>>>  
 -----  
 Zero Order Coef -310.49  
 First Order Coef 2802.82  
 Second Order Coef -7.70  
 -----

<<<<< 9um >>>>>  
 -----  
 -315.38  
 1345.11  
 4.12  
 -----

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	-0.000	0.0000
0.040	0.040	0.0003
0.107	0.108	-0.0013
0.147	0.146	0.0011
0.285	0.285	-0.0001

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	0.000	-0.0000
0.040	0.040	0.0003
0.107	0.108	-0.0012
0.147	0.146	0.0010
0.285	0.285	-0.0001

<<<<< 3um >>>>>  
 -----

<<<<< 9um >>>>>  
 -----

Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1

Sample	3um	9um
Sample #1	3160.00	3206.00
Sample #2	3211.00	3231.00
Sample #3	3128.00	3212.00
Sample #4	3186.00	3247.00
Avg	3175.0000	3230.0000
STD DEV	42.5793	17.5214
REL STD DEV	1.341	0.542
H2O adjust (mg/l*10k)	634	579

Atmospheric Pressure = 1018

\*\*\*\*\*CALIBRATION SUCCESSFUL\*\*\*\*\*

*Calibration  
 Deb Kashner  
 P. 2 of 2*

*80-004201  
 P. 4 of 21*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004201  
Location = TOXL      8164.13.00 06/09  
~~12/23/2009~~ 12/22/2009 DK      14:00  
\*\*Internal Test Procedure Calibration\*\*  
Target: 080 ReloadOffset:6327

*ITP calibration  
Deb Kashur*

*80-004201  
P. S. 12/21/09*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004201  
Location = TOXL      8164.13.00 06/09  
~~12/23/2009~~      14:01  
12/22/2009 DK

ITP CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:01
02 ITP Check	0.081	14:02
03 Room Air	0.000	14:02

Deb Kashur  
Operator Signature  
DEB KASHUR

Remarks:

*ITP check after calibration*

Form 106-I8000

*80-004201  
P. 04/21/02*

Intoxilyzer Test Record and Checklist  
NDDAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer  
North Dakota Model 8000 SN 80-004201  
Location = 70XL 8164.13.00 06/09  
12/22/2009 14:11

\*\*\*\*\* Printer Test \*\*\*\*\*

12 char/line  
abcdefghijkl  
ABCDEFGHIJKL

16 char per line  
abcdefghijklmnop  
ABCDEFGHIJKLMN

21 chars per line  
abcdefghijklmnopqrstu  
ABCDEFGHIJKLMNQRSTU

24 characters per line  
abcdefghijklmnopqrstuwx  
ABCDEFGHIJKLMNQRSTUWX

32 characters per line  
abcdefghijklmnopqrstuwx123456  
ABCDEFGHIJKLMNQRSTUWXYZ123456

42 characters per line  
abcdefghijklmnopqrstuwx1234567890abcdef  
ABCDEFGHIJKLMNQRSTUWXYZ1234567890ABCDEF

Current Instrument Setup  
Data Entry Mode: Enabled  
Start Test Sequence: 0ABACABA  
Display Prelim Res?: Yes  
Display Third Digit? Yes  
Inhib Printer(Y/N)? No  
Display Volume? No  
Disable On MemFull? Yes  
# of Print Copies? 1  
Select Std (0/4/1)? Dry  
Standard Value? 0.000  
Standard Lot #? 659358  
Standard Cyl #? 17  
Standard Expiration? 09/03/2011  
Oper No? 777777

Flow Cal. Date: 06/03/2009  
Slope 696  
Intercept -646527

IR Calibration Date: 12/23/2009  
3um 9um

	3um	9um
0th Coef(*100):	-31049	-31538
1st Coef(*100):	280282	134511
2nd Coef(*100):	-770	412
H2O adj(ng/l*10k):	634	579

\*\*\*\*\* Test End \*\*\*\*\*

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Print test*

Form 106-I8000

80-004201  
A747  
21

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004201  
Location = TOXL                    8164.13.00 06/09  
12/22/2009                            14:09

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            659358  
Standard Cyl #?            17  
Standard Expiration?        09/03/2011  
Oper No?                    777777

Flow Cal. Date:            06/03/2009  
Slope                      696  
Intercept                  -646527

IR Calibration Date:        12/23/2009

	3um	9um
0th Coef(*100):	-31049	-31538
1st Coef(*100):	280282	134511
2nd Coef(*100):	-770	412
H2O adj(mg/l*10k):	634	579

\*\*\*\*\* Printer Test End \*\*\*\*\*

Deb Kasher

Operator Signature  
DEB KASHUR

Remarks:

Print test



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004201  
Location = TOXL                    8164.13.00 06/09  
12/22/2009                            14:24

Test	AC	Time
01 Room Air	0.000	14:25
02 *Subject Test	REF*	14:25
03 Room Air	0.000	14:25

\*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 659358  
Cyl No = 17  
Expiration Date = 09/03/2011  
County = 08                            Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Camera/Refusal check*

Form 106-I8000

80-004201  
P. 9 of 21/22



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004201  
Location = TOXL      8164.13.00 06/09  
12/22/2009      14:30

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:31
02 Std. Sol.	0.000	14:35
03 Room Air	0.000	14:36
04 Std. Sol.	0.000	14:36
05 Room Air	0.000	14:37
06 Std. Sol.	0.000	14:37
07 Room Air	0.000	14:38

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*zero alcohol      0.000 AC*

Form 106-I8000

*80-004201  
P. 11/21/09*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004201  
Location = TOXL      8164.13.00 06/09  
12/22/2009      14:41

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:42
02 Std. Sol.	0.019	14:42
03 Room Air	0.000	14:43
04 Std. Sol.	0.019	14:44
05 Room Air	0.000	14:44
06 Std. Sol.	0.019	14:45
07 Room Air	0.000	14:45

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kasher

Operator Signature  
DEB KASHUR

Remarks:

*Low AC      0.019 AC*

Form 106-I8000

*80-004201  
P.12 of 21  
VM*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004201  
Location = TOXL      8164.13.00 06/09  
12/22/2009      14:53

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:54
02 Std. Sol.	0.285	14:55
03 Room Air	0.000	14:55
04 Std. Sol.	0.286	14:56
05 Room Air	0.000	14:57
06 Std. Sol.	0.285	14:57
07 Room Air	0.000	14:58

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Linearity test 0.285 AC*

Form 106-I8000

80-004201  
P. 13 of 21  
1/12

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004201  
Location = TOXL      8164.13.00 06/09  
12/22/2009      15:04

Test	AC	Time
01 Room Air	0.000	15:05
02 *Subject Test	INT*	15:05
03 Room Air	0.000	15:06

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Simul Ser No = 00000  
Std Sol No = 000  
County = 08      Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*acetone interferent*

Form 106-I8000

80-004201  
P. 14 of 21

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004201  
Location = TOXL      8164.13.00 06/09  
12/22/2009      15:06

Test	AC	Time
01 Room Air	0.000	15:07
02 *Subject Test	INT*	15:07
03 Room Air	0.000	15:08

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Simul Ser No = 00000  
Std Sol No = 000  
County = 08      Oper No. = 777777

Deb Kasher

Operator Signature  
DEB KASHUR

Remarks:

*Methanol interferent*

Form 106-I8000

80-004201  
P. 15 of 21 ym

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004201  
Location = TOXL      8164.13.00 06/09  
12/22/2009      15:59

Test	AC	Time
01 Room Air	0.000	16:00
02 *Subject Test	INT*	16:00
03 Room Air	0.000	16:01

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Simul Ser No = 00000  
Std Sol No = 000  
County = 08      Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*IPA interferent*

Form 106-I8000

*80-004201  
P 16 of 21 1/24*



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004201  
Location = TOXL                      8164.13.00 06/09  
12/22/2009                              16:03

Test	AC	Time
01 Room Air	AMB*	16:04
02 Room Air	0.000	16:05

\*Invalid Test  
Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                      Weight = 150  
Test = DUI                              Cit = NA  
Dr. Lic. = ND/DIS921456  
Simul Ser No = 00000  
Std Sol No = 000  
County = 08                              Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*ambient test*

Form 106-I8000

80-004201  
P. 17-21 YH

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004201  
Location = TOXL                    8164.13.00 06/09  
12/22/2009                            16:07

Test	AC	Time
01 Room Air	0.000	16:08
02 Invalid Sample	X.XXX	16:08
03 Room Air	0.000	16:09

\*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Simul Ser No = 00000  
Std Sol No = 000  
County = 08                            Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Invalid sample test*

Form 106-I8000

*80-004201  
P. 18 of 21/22*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004201  
Location = TOXL                    8164.13.00 06/09  
12/22/2009                            16:21

Test	AC	Time
01 Diagnostic	OK	16:22
02 Room Air	RFI*	16:22
03 Room Air	0.000	16:22

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 659358  
Cyl No = 17  
Expiration Date = 09/03/2011  
County = 08                            Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kashur  
Operator Signature  
DEB KASHUR

Remarks:

*RFI test*

Form 106-I8000

*80-004201  
P. 19 of 21 KM*

Intoxilyzer Test Record and Checklist  
 NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
 North Dakota Model 8000      SN 80-004201  
 Location = TOXL      8164.13.00 06/09  
 12/22/2009      16:32

Test	AC	Time
01 Room Air	0.000	16:33
02 Std. Gas	0.080	16:33
03 Room Air	0.000	16:34
04 Std. Gas	0.079	16:34
05 Room Air	0.000	16:35
06 Std. Gas	0.079	16:35
07 Room Air	0.000	16:35
08 Std. Gas	0.079	16:36
09 Room Air	0.000	16:36
10 Std. Gas	0.079	16:37
11 Room Air	0.000	16:37
12 Std. Gas	0.079	16:38
13 Room Air	0.000	16:38
14 Std. Gas	0.079	16:38
15 Room Air	0.000	16:39
16 Std. Gas	0.080	16:39
17 Room Air	0.000	16:40
18 Std. Gas	0.079	16:40
19 Room Air	0.000	16:41
20 Std. Gas	0.080	16:41
21 Room Air	0.000	16:42
22 Std. Gas	0.081	16:42
23 Room Air	0.000	16:42
24 Std. Gas	0.080	16:43
25 Room Air	0.000	16:43
26 Std. Gas	0.080	16:44
27 Room Air	0.000	16:44
28 Std. Gas	0.080	16:45
29 Room Air	0.000	16:45
30 Std. Gas	0.079	16:45
31 Room Air	0.000	16:46
32 Std. Gas	0.079	16:46
33 Room Air	0.000	16:47
34 Std. Gas	0.080	16:47
35 Room Air	0.000	16:48
36 Std. Gas	0.080	16:48
37 Room Air	0.000	16:49
38 Std. Gas	0.080	16:49
39 Room Air	0.000	16:49
40 Std. Gas	0.080	16:50
41 Room Air	0.000	16:50
42 Cal Check Stats:		
43 Average	0.0796	
44 Std Dev	0.0006	
45 Rel Std Dev	0.76%	

*Stability test  
 Deb Kashner  
 p. 1 of 2*

*80-004201  
 P. 20 of 21 FN*

Operator Signature  
DEB KASHUR

Remarks:

Lot #: 659358 Cyl #: 17 Exp. date: 03 Sept 2011

Form 106-I8000

Stability test  
Deb Kashur  
p. 2082

80-004201  
P. 21/21 KH