



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>William Petty</u>		
Location: <u>GF AF</u>		
Intoxilyzer® Serial Number: <u>80-002667</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>39 DK</u>	Expiration Date: <u>09/03/11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.080</u> %		
Test 2: <u>0.081</u> %		
Test 3: <u>0.081</u> %		
(ATTACH TEST RECORD)		

Will. Petty
Chemical Test Operator Signature

22 Dec 10
Date

Deb Kashner
Reviewed By

22 Mar 2011
Date

80-002667
P. 84/85
jam


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = GFAF 8164.13.00 06/09
12/22/2010 09:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:23
02 Std. Gas	0.080	09:23
03 Room Air	0.000	09:24
04 Std. Gas	0.081	09:24
05 Room Air	0.000	09:25
06 Std. Gas	0.081	09:25
07 Room Air	0.000	09:26

Lot No = 659358
Cyl No = 39
Exp Date = 09/03/2011
County = 18 Oper No. = 120011



Operator Signature
WILLIAM R. PETTY

Remarks: *ACA Test*

Form 106-I8000

*80-002667
P. 85 of 85
RM*