



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>MICHAEL McCLOUD</u>		
Location: <u>WARD</u>		
Intoxilyzer® Serial Number: <u>80-003066</u>		
Gas Lot Number: <u>654358</u>	Gas Cylinder Number: <u>32</u>	Expiration Date: <u>09/03/11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.081</u> %		
Test 2: <u>0.081</u> %		
Test 3: <u>0.080</u> %		
(ATTACH TEST RECORD)		

Michael McCLOUD
Chemical Test Operator Signature

12/23/10
Date

Deb Kashner
Reviewed By

10 Jan 2011
Date

80-003066
P. 121 of 122
FM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003066
Location = WARD 8164.13.00 06/09
12/23/2010 15:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:10
02 Std. Gas	0.081	15:11
03 Room Air	0.000	15:11
04 Std. Gas	0.081	15:12
05 Room Air	0.000	15:12
06 Std. Gas	0.080	15:13
07 Room Air	0.000	15:13

Lot No = 659358
Cyl No = 32
Exp Date = 09/03/2011
County = 51 Oper No. = 049511


Operator Signature
MICHAEL MCCLOUD

Remarks:

Form 106-18000

80-003066
P. 120 of 120
JFM