



## Intoxilyzer 8000 Initial Inspection

<b>Serial Number</b>	<b>80-003061</b>
----------------------	------------------

- A. Power On  5 min. countdown  1 min.
- B. Security Checks.  Level 1  Level 2  Level 3
- C. Gas Cylinder Setup
- D. Printer. Level 1 Func P  Paper Feed  Internal  External
- E. Display is working.
- F. Configure Start Test Switch. Level 3 Func S Sub T
- G. Setup Preliminary Data. Level 2 Func E
- H. Configure General Setup. Level 3 Func S Sub G.
- I. Configure Port. Level 3 Func S Sub N
- J. Check Camera. ABA Mode  Oper ID  ND Lic.  Hot Key "R"
- K. Diagnostic Test. Level 2 Func D
- L. Check Ser Num / Version Level 2 Func V  
80-003061 Serial No. 8164.13 Version 1362.14 Display 7 CPLD
- M. Detector Voltages. Level 3 Func D Sub D  
3 12857 s 2 C 47.00  
9 13008 s 2 B 45 0 RFI Ind.
- N. Hot Keys  
 F5 Tone  
 F6 Green LED  
 F7 Red LED  
 F9 Fan  
 F10 Pump  
 F11 Sol Output
- O. RFI Dectector
- P. Atmospheric Monitor. Level 3 Func D Sub A. P 1018 CF 0.99

08/09

80-003061  
P. 10/21  
FM

OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION  
 Toxicology Section/Breath Alcohol Program  
 Instrument Certification and Maintenance Unit

BrF-005-8

- ✓ Q. Flow Monitor. Level 3 Func D Sub F. R 189 F 0.000 L/s
- ✓ R. Temperature Monitor. Level 3 Func D Sub T  
 C 47.0 A 48.0 I 0.0  
 B 45.0 V 0.0
- ✓ S. Gas Tank Monitor. Level 3 Func D Sub G 64 psi Intox 60 psi Reg
- ✓ T. Configure Wet Bath Sim. Level 1 Func S
- ✓ U. Zero Alcohol Test Level 1 Func C  
DR3409 Sim SN --- Lot 0.000 Rep Value  
0.000 AC 0.000 AC 0.000 AC
- ✓ V. Low AC Test. Level 1 Func C  
~~DK-DR-3~~ DR5144 Sim SN --- Lot 0.019 Rep Value  
0.019 AC 0.019 AC 0.019 AC
- ✓ W. Linearity Test. Level 1 Func C  
DR5187 Sim SN --- Lot 0.285 Rep Value  
0.285 AC 0.286 AC 0.287 AC
- ✓ X. Interferent Test. Level 1 Func B  
 Acetone DR3845 Sim SN --- Lot 0.05% Conc  
 ✓ Display ✓ Printout  
 Methanol DR10573 Sim SN --- Lot 0.10% Conc  
 ✓ Display ✓ Printout  
 IPA DR5111 Sim SN --- Lot 0.10% Conc  
 ✓ Display ✓ Printout
- ✓ Y. Ambient Test  
~~NA~~ Display ~~NA~~ Printout
- ✓ Z. Invalid Sample Test  
 ✓ Display ✓ Printout
- ✓ AA. RFI Detector Test. CMS Mode  
 ✓ Display ✓ Printout
- ✓ BB. Configure Dry Gas Standard. Level 1 Func S
- ✓ CC. Stability Test at 0.080AC. Level 3 Func T Sub S  
0.0808 Ave. 0.0004 Std Dev 0.56% Rel Std Dev
- ✓ DD. Maintenance/Repairs Software upgrade 8/6/13. Calibration of optical bench & ITP.

24 ~~23~~ Dec 2009  
 Date  
30 Dec 2009  
 Date

Deb Kashur  
 Signature  
Mary Plason  
 Reviewed by

TOXL  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-003061  
 12/23/2009 10:45:58

Auto Calibration

pg 1 of 2

<<<<< 3um >>>>> <<<<< 9um >>>>>

-----  
 Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1  

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	0.0710	(0.0220)	0.1260	(-0.0030)
Sample #2	0.0660	(0.0660)	0.1070	(0.0020)
Sample #3	0.0820	(0.0810)	0.1270	(0.0190)
Sample #4	0.0770	(0.1010)	0.1020	(0.0150)
Avg % Abs	0.0750	(0.0827)	0.1120	(0.0120)
STD DEV	0.0082	(0.0176)	0.0132	(0.0089)
REL STD DEV	10.914	(21.241)	11.811	(74.068)

-----  
 Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1  

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	0.8370	(-0.0260)	1.5470	(-0.0310)
Sample #2	0.7740	(0.0220)	1.5080	(0.0140)
Sample #3	0.8200	(0.0140)	1.4810	(0.0280)
Sample #4	0.7700	(0.0310)	1.4730	(0.0260)
Avg % Abs	0.7880	(0.0223)	1.4873	(0.0227)
STD DEV	0.0278	(0.0085)	0.0183	(0.0076)
REL STD DEV	3.526	(38.082)	1.233	(33.405)

-----  
 Solution = 0.107 g/210L or 0.5095 mg/l, Samples = 4, Discarded = 1  

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	2.1200	(-0.0260)	3.8930	(-0.0150)
Sample #2	2.0530	(0.0350)	3.8130	(0.0500)
Sample #3	1.9940	(0.0610)	3.7800	(0.0630)
Sample #4	2.0290	(0.0490)	3.7930	(0.0470)
Avg % Abs	2.0253	(0.0483)	3.7953	(0.0533)
STD DEV	0.0297	(0.0130)	0.0166	(0.0085)
REL STD DEV	1.465	(26.923)	0.438	(15.947)

-----  
 Solution = 0.147 g/210L or 0.7000 mg/l, Samples = 4, Discarded = 1  

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	2.7450	(0.0020)	5.1680	(0.0020)
Sample #2	2.6780	(0.0770)	5.0360	(0.1110)
Sample #3	2.6950	(0.0680)	5.0740	(0.0760)
Sample #4	2.6700	(0.0690)	5.0420	(0.0840)
Avg % Abs	2.6810	(0.0713)	5.0507	(0.0903)
STD DEV	0.0128	(0.0049)	0.0204	(0.0183)
REL STD DEV	0.476	(6.915)	0.404	(20.302)

*Calibration  
 Deb Koshur  
 p. 1 of 2*

-----  
 Solution = 0.285 g/210L or 1.3571 mg/l, Samples = 4, Discarded = 1  

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	5.1870	(0.0110)	9.6160	(0.0080)
Sample #2	5.0930	(0.0890)	9.4910	(0.1280)
Sample #3	5.1300	(0.0700)	9.5410	(0.1040)
Sample #4	5.0760	(0.1110)	9.5050	(0.1290)
Avg % Abs	5.0997	(0.0900)	9.5123	(0.1203)
STD DEV	0.0276	(0.0205)	0.0258	(0.0142)
REL STD DEV	0.541	(22.798)	0.271	(11.762)

*80-003061  
 P. 3 of 21  
 1/21*

TOXL  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-003061  
 12/23/2009 10:45:58

Auto Calibration

pg 2 of 2

<<<<< 3um >>>>>

<<<<< 9um >>>>>

-----  
 Zero Order Coef -192.13  
 First Order Coef 2615.28  
 Second Order Coef 16.60  
 -----

-----  
 -155.60  
 1366.94  
 8.05  
 -----

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	0.000	-0.0001
0.040	0.039	0.0005
0.107	0.109	-0.0016
0.147	0.146	0.0013
0.285	0.285	-0.0001

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	-0.000	0.0001
0.040	0.040	0.0002
0.107	0.108	-0.0011
0.147	0.146	0.0010
0.285	0.285	-0.0001

<<<<< 3um >>>>>

<<<<< 9um >>>>>

-----  
 Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1  
 Sample

Sample #1	3241.00	3329.00
Sample #2	3281.00	3368.00
Sample #3	3321.00	3383.00
Sample #4	3202.00	3364.00
Avg	3268.0000	3371.6667
STD DEV	60.5558	10.0167
REL STD DEV	1.853	0.297
H2O adjust (mg/l*10k)	541	438

Atmospheric Pressure = 1020

\*\*\*\*\*CALIBRATION SUCCESSFUL\*\*\*\*\*

*Calibration  
 Deb Kashner  
 P. 2 of 2*

80-003061  
 P. 4 of 21

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003061  
Location = TOXL      8164.13.00 06/09  
12/23/2009      11:41  
\*\*Internal Test Procedure Calibration\*\*  
Target: 080 ReloadOffset:5765

*ITP Calibration  
Deb Kashner*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003061  
Location = TOXL      8164.13.00 06/09  
12/23/2009      11:42

ITP CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:42
02 ITP Check	0.080	11:43
03 Room Air	0.000	11:43

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:

ITP check

Form 106-I8000

80-003061  
P. 6/21/09

TOXL  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-003061  
12/23/2009 10:45:59

Auto Calibration  
Max Power Res Value = 22  
Auto Range Res Value = 13  
Intoxilyzer Test Record and Checklist  
NDDAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer  
North Dakota Model 8000 SN 80-003061  
Location = TOXL 8164.13.00 06/09  
12/23/2009 11:50

\*\*\*\*\* Printer Test \*\*\*\*\*

12 char/line  
abcdefghijkl  
ABCDEFGHIJKL

16 char per line  
abcdefghijklmnop  
ABCDEFGHIJKLMN

21 chars per line  
abcdefghijklmnopqrstu  
ABCDEFGHIJKLMNQRSTU

24 characters per line  
abcdefghijklmnopqrstuvw  
ABCDEFGHIJKLMNQRSTU

32 characters per line  
abcdefghijklmnopqrstuvwxy123456  
ABCDEFGHIJKLMNQRSTUWXYZ123456

42 characters per line  
abcdefghijklmnopqrstuvwxy1234567890abcdef  
ABCDEFGHIJKLMNQRSTUWXYZ1234567890ABCDEF

Current Instrument Setup  
Data Entry Mode: Enabled  
Start Test Sequence: DABACABA  
Display Prelim Rslt? Yes  
Display Third Digit? Yes  
Inhib Printer(Y/N)? No  
Display Volume? No  
Disable On Memfull? Yes  
# of Print Copies? 1  
Select Std (D/W/I)? Wet  
Standard Value? 0.000  
Standard Solution #? 000  
Sin Serial #? 00000  
Oper No? 77777

Flow Cal. Date: 06/15/2009  
Slope: 737  
Intercept: -507305

IR Calibration Date: 12/23/2009  
3um 9um  
-----  
0th Coef(\*100): -19213 -15560  
1st Coef(\*100): 261528 136694  
2nd Coef(\*100): 1660 805  
H2O adj(ng/l\*10K): 541 438

\*\*\*\*\* Test End \*\*\*\*\*

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Print Test*

Form 106-18000

80-003061  
P. 7 of 21 YAK

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer           Alcohol Analyzer  
North Dakota Model 8000       SN 80-003061  
Location = TOXL                8164.13.00 06/09  
12/23/2009                      11:51

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?         Yes  
Inhib Printer(Y/N)?         No  
Display Volume?             No  
Disable On Memfull?         Yes  
# of Print Copies?           1  
Select Std (D/W/I)?         Wet  
Standard Value?             0.000  
Standard Solution #?         000  
Sim Serial #?                00000  
Oper No?                     777777

Flow Cal. Date:             06/15/2009  
Slope                        737  
Intercept                    -507305

IR Calibration Date:        12/23/2009  
                              3um                    9um

-----  
0th Coef(\*100):            -19213           -15560  
1st Coef(\*100):            261528           136694  
2nd Coef(\*100):            1660                805  
H2O adj(mg/l\*10k):         541                438

\*\*\*\*\* Printer Test End \*\*\*\*\*

          Deb Kasher            
Operator Signature  
DEB KASHUR

Remarks:

Print Test

Form 106-I8000

80-003061  
P. Page 21 of 22



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-003061  
Location = TOXL                    8164.13.00 06/09  
12/23/2009                            14:01

Test	AC	Time
01 Room Air	0.000	14:02
02 *Subject Test	REF*	14:02
03 Room Air	0.000	14:02

\*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Simul Ser No = 00000  
Std Sol No = 000  
County = 08                            Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Camera/Refusal check*

Form 106-I8000

80-003061  
P. 9 of 21 vcm

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003061  
Location = TOXL      8164.13.00 06/09  
12/23/2009      14:03

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Diagnostic test*

Form 106-I8000

80-003061  
P. 16 of 21 JLN

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003061  
Location = TOXL      8164.13.00 06/09  
12/23/2009      11:45

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:45
02 Std. Sol.	0.000	11:46
03 Room Air	0.000	11:47
04 Std. Sol.	0.000	11:47
05 Room Air	0.000	11:48
06 Std. Sol.	0.000	11:48
07 Room Air	0.000	11:49

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

zero alcohol 0.000AC

Form 106-18000

80-003061  
P.11 of 21 JCN

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003061  
Location = TOXL      8164.13.00 06/09  
12/23/2009      14:07

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:07
02 Std. Sol.	0.019	14:08
03 Room Air	0.000	14:09
04 Std. Sol.	0.019	14:10
05 Room Air	0.000	14:10
06 Std. Sol.	0.019	14:11
07 Room Air	0.000	14:12

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:

Low AC 0.019 AC

Form 106-I8000

80-003061  
P. 12 of 21 ycm

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003061  
Location = TOXL      8164.13.00 06/09  
12/23/2009      16:36

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:37
02 Std. Sol.	0.285	16:38
03 Room Air	0.000	16:38
04 Std. Sol.	0.286	16:39
05 Room Air	0.000	16:40
06 Std. Sol.	0.287	16:41
07 Room Air	0.000	16:41

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

Linearity test      0.285 AC

Form 106-I8000

80-003061  
P. 13 of 21 VFL

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-003061  
Location = TOXL                    8164.13.00 06/09  
12/24/2009                            08:20

Test	AC	Time
01 Room Air	0.000	08:20
02 *Subject Test	INT*	08:21
03 Room Air	0.000	08:21

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 659358  
Cyl No = 17  
Expiration Date = 09/03/2011  
County = 08                            Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Acetone interferent*

Form 106-I8000

80-003061  
P. 14 of 21 FA

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003061  
Location = TOXL                    8164.13.00 06/09  
12/24/2009                            08:22

Test	AC	Time
01 Room Air	0.000	08:22
02 *Subject Test	INT*	08:23
03 Room Air	0.000	08:23

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 659358  
Cyl No = 17  
Expiration Date = 09/03/2011  
County = 08                            Oper No. = 777777

Deb Kasher

Operator Signature  
DEB KASHUR

Remarks:

*Methanol interferent*

Form 106-I8000

80-003061  
P. 15 of 21 FH


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003061  
Location = TOXL      8164.13.00 06/09  
12/24/2009      08:23

Test	AC	Time
01 Room Air	0.000	08:24
02 *Subject Test	INT*	08:24
03 Room Air	0.000	08:25

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 659358  
Cyl No = 17  
Expiration Date = 09/03/2011  
County = 08      Oper No. = 777777



Operator Signature  
DEB KASHUR

Remarks:

*IPA interferent*

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-003061  
Location = TOXL                    8164.13.00 06/09  
12/24/2009                            08:28

Test	AC	Time
01 Room Air	AMB*	08:28
02 Room Air	0.000	08:29

\*Invalid Test  
Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 659358  
Cyl No = 17  
Expiration Date = 09/03/2011  
County = 08                            Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Ambient Test*

Form 106-I8000

80-003061  
P-12/21/09



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003061  
Location = TOXL      8164.13.00 06/09  
12/24/2009      08:34

Test	AC	Time
01 Diagnostic	OK	08:36
02 Room Air	RFI*	08:36
03 Room Air	0.000	08:36

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 659358  
Cyl No = 17  
Expiration Date = 09/03/2011  
County = 08      Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kasher

Operator Signature  
DEB KASHUR

Remarks:

RFI test

Form 106-18000

80-003061  
P. 19 of 21 VAK

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003061  
Location = TOXL      8164.13.00 06/09  
12/24/2009      08:54

Test	AC	Time
01 Room Air	0.000	08:55
02 Std. Gas	0.081	08:55
03 Room Air	0.000	08:56
04 Std. Gas	0.081	08:56
05 Room Air	0.000	08:57
06 Std. Gas	0.080	08:57
07 Room Air	0.000	08:58
08 Std. Gas	0.081	08:58
09 Room Air	0.000	08:59
10 Std. Gas	0.081	08:59
11 Room Air	0.000	08:59
12 Std. Gas	0.081	09:00
13 Room Air	0.000	09:00
14 Std. Gas	0.081	09:01
15 Room Air	0.000	09:01
16 Std. Gas	0.081	09:02
17 Room Air	0.000	09:02
18 Std. Gas	0.081	09:02
19 Room Air	0.000	09:03
20 Std. Gas	0.081	09:03
21 Room Air	0.000	09:04
22 Std. Gas	0.080	09:04
23 Room Air	0.000	09:05
24 Std. Gas	0.081	09:05
25 Room Air	0.000	09:06
26 Std. Gas	0.081	09:06
27 Room Air	0.000	09:07
28 Std. Gas	0.081	09:07
29 Room Air	0.000	09:08
30 Std. Gas	0.081	09:08
31 Room Air	0.000	09:08
32 Std. Gas	0.081	09:09
33 Room Air	0.000	09:09
34 Std. Gas	0.080	09:10
35 Room Air	0.000	09:10
36 Std. Gas	0.081	09:11
37 Room Air	0.000	09:11
38 Std. Gas	0.080	09:12
39 Room Air	0.000	09:12
40 Std. Gas	0.080	09:12
41 Room Air	0.000	09:13
42 Cal Check Stats:		
43 Average	0.0808	
44 Std Dev	0.0004	
45 Rel Std Dev	0.56%	

*Stability test  
Deb Kashner  
p. 1 of 2*

*80-003061  
P. 20 of 21 ym*

Operator Signature  
DEB KASHUR

Remarks:

Lot #: 659358

Cyl #: 17

Exp. date: 03 Sept 2011

Form 106-I8000

Stability test  
Deb Kashur  
p. 2 of 2

80-003661  
P. 21 of 21 YAM