

INTOXILYZER 5000 G and 8000 INSPECTION

Serial No.: 80-003061

Location: TOXL

- |  | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes?   | <u>✓</u>        |
| B. Is a three-pronged grounded outlet used?  | <u>✓</u>        |
| C. Is the breath tube heated?  | <u>✓</u>        |
| D. Is the diagnostic check complete?   | <u>✓</u>        |
| E. Is the time, date, and year correct (re-set if necessary)?  | <u>✓</u>        |
| F. Print test. (Attach test record.)   | <u>✓</u>        |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5144</u> Lot No. <u>    </u> Rep AC <u>0.019</u>   | <u>✓</u>        |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5187</u> Lot No. <u>    </u> Rep AC <u>0.285</u>                                | <u>✓</u>        |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.<br>(Attach test record.)<br>Sim SN <u>DR5114</u> Lot No. <u>    </u> Rep AC <u>0.100</u> | <u>✓</u>        |
| J. RFI check. Run CMS mode. Key radio on first room air.<br>(Attach test record.)  | <u>✓</u>        |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.<br>(Attach test record.)  |                 |

Lot No. 659358 Cyl No. 17 Exp. Date 03 Sept 2011

Test 1 0.079 Test 4 0.081 Test 7 0.081

Test 2 0.080 Test 5 0.080 Test 8 0.080

Test 3 0.080 Test 6 0.080 Test 9 0.081

Average                     

L. Remarks/Maintenance record: Software upgrade 8164.B. Annual inspection.

<sup>24</sup>  
~~23~~ Dec 2009  
Date

Deb Kasher  
Field Inspector's Signature

31 Dec 2009  
Date

Mary Blain  
Reviewed by

80-003061  
P. 1 of 9  
JAN

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003061  
Location = TOXL                8164.13.00 06/09  
12/23/2009                      11:51

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Wet  
Standard Value?            0.000  
Standard Solution #?        000  
Sim Serial #?              00000  
Oper No?                    777777

Flow Cal. Date:            06/15/2009  
Slope                      737  
Intercept                  -507305

IR Calibration Date:        12/23/2009  
                                  3um                            9um

-----  
0th Coef(\*100):            -19213                        -15560  
1st Coef(\*100):            261528                        136694  
2nd Coef(\*100):            1660                            805  
H2O adj (mg/l\*10k):        541                            438

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
DEB KASHUR

Remarks:

*Print test*

Form 106-I8000

80-003061  
P. 2 of 9 ysk

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003061  
Location = TOXL      8164.13.00 06/09  
12/23/2009      14:12

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:12
02 Std. Sol.	0.019	14:13
03 Room Air	0.000	14:14
04 Std. Sol.	0.020	14:15
05 Room Air	0.000	14:15
06 Std. Sol.	0.019	14:16
07 Room Air	0.000	14:16

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kashur  
Operator Signature  
DEB KASHUR

Remarks:

Low AC      0.019 AC

Form 106-I8000

80-003061  
P. 3 of 9

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003061  
Location = TOXL      8164.13.00 06/09  
12/24/2009      08:13

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:14
02 Std. Sol.	0.286	08:14
03 Room Air	0.000	08:15
04 Std. Sol.	0.286	08:16
05 Room Air	0.000	08:17
06 Std. Sol.	0.286	08:17
07 Room Air	0.000	08:18

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Linearity test*

Form 106-18000

80-003061  
p. 4 of 9 JKA

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-003061  
Location = TOXL                    8164.13.00 06/09  
12/24/2009                            08:25

Test	AC	Time
01 Room Air	0.000	08:26
02 *Subject Test	INT*	08:26
03 Room Air	0.000	08:27


\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 659358  
Cyl No = 17  
Expiration Date = 09/03/2011  
County = 08                            Oper No. = 777777



Operator Signature  
DEB KASHUR

Remarks:



Form 106-I8000

80-003061  
P. Sof 9/24

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003061  
Location = TOXL                8164.13.00 06/09  
12/24/2009                        08:36

Test	AC	Time
01 Diagnostic	OK	08:37
02 Room Air	RFI*	08:37
03 Room Air	0.000	08:38

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                Weight = 150  
Test = DUI                        Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 659358  
Cyl No = 17  
Expiration Date = 09/03/2011  
County = 08                        Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

*RFI test*

Form 106-18000

80-003061  
P. 6 of 9 JAL

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003061  
Location = TOXL      8164.13.00 06/09  
12/24/2009      08:42

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:43
02 Std. Gas	0.079	08:43
03 Room Air	0.000	08:44
04 Std. Gas	0.080	08:44
05 Room Air	0.000	08:45
06 Std. Gas	0.080	08:45
07 Room Air	0.000	08:46

Lot No = 659358  
Cyl No = 17  
Exp Date = 09/03/2011  
County = 08      Oper No. = 777777

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:

*Calibration check*  
Form 106-18000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003061  
Location = TOXL      8164.13.00 06/09  
12/24/2009      08:46

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:47
02 Std. Gas	0.081	08:47
03 Room Air	0.000	08:48
04 Std. Gas	0.080	08:48
05 Room Air	0.000	08:48
06 Std. Gas	0.080	08:49
07 Room Air	0.000	08:49

Lot No = 659358  
Cyl No = 17  
Exp Date = 09/03/2011  
County = 08      Oper No. = 777777

Deb Kashur  
Operator Signature  
DEB KASHUR

Remarks:

*Calibration check*

Form 106-18000

80-003061  
P. 8 of 9 PM



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003061  
Location = TOXL      8164.13.00 06/09  
12/24/2009      08:50

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:50
02 Std. Gas	0.081	08:50
03 Room Air	0.000	08:51
04 Std. Gas	0.080	08:51
05 Room Air	0.000	08:52
06 Std. Gas	0.081	08:52
07 Room Air	0.000	08:53

Lot No = 659358  
Cyl No = 17  
Exp Date = 09/03/2011  
County = 08      Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check*

Form 106-I8000

*80-003061  
P. 9 of 9 YAM*