



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59281 (5-2011)

Serial Number 80-004194	Instrument Location Langdon - Cavalier County SO
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Properly Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. RFI Test (Use CMS Mode; Key Radio During First Room Air).
- 7. Repair and/or Maintenance Performed (if Any): N/A
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer®8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>[Signature]</i>	Date 12/16/2011 12/27/12 ^{DR}
Reviewed By (Crime Laboratory Use Only) <i>Deb Kashner</i>	Date 02 Apr 2012

This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Certified By: *[Signature]* Certified Date: 08 Nov. 16

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004194
Location = LANG 8164.13.00 06/09
12/16/2011 14:18

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 23411080A1
Standard Cyl #? 04
Standard Expiration? 10/01/2013
Oper No? 067518

Flow Cal. Date: 06/23/2009
Slope 681
Intercept -622430

IR Calibration Date: 06/29/2011

	3um	9um
0th Coef(*100):	1196	-615
1st Coef(*100):	247548	123137
2nd Coef(*100):	7878	2577
H2O adj(mg/l*10k):	583	655

***** Printer Test End *****



Operator Signature
DAVID J. ZEIS

Remarks:


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004194
Location = LANG 8164.13.00 06/09
12/16/2011 14:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:28
02 Std. Gas	0.082	14:28
03 Room Air	0.000	14:29
04 Std. Gas	0.082	14:29
05 Room Air	0.000	14:30
06 Std. Gas	0.082	14:30
07 Room Air	0.000	14:31

Lot No = 23411080A1
Cyl No = 04
Exp Date = 10/01/2013
County = 10 Oper No. = 067518


Operator Signature
DAVID J. ZEIS

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004194
Location = LANG 8164.13.00 06/09
12/16/2011 14:34

Test	AC	Time
01 Room Air	RFI*	14:39
02 Room Air	0.000	14:39

*Invalid Test
Inhibited - RFI

Sub Name = ZE, DA J
Sub DOB = 01/01/1954
Sub Sex = Male Weight = 200
Test = DUI Cit = 1111
Dr. Lic. = ND/ZEI201111
Lot No = 23411080A1
Cyl No = 04
Expiration Date = 10/01/2013
County = 10 Oper No. = 067518



Operator Signature
DAVID J. ZEIS

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004194
Location = LANG 8164.13.00 06/09
12/27/2011 10:07

Test	AC	Time
01 Diagnostic	OK	10:10
02 Room Air	RFI*	10:10
03 Room Air	0.000	10:10

*Invalid Test
Inhibited - RFI

Sub Name = ZE, DA J
Sub DOB = 01/01/1954
Sub Sex = Male Weight = 200
Test = DUI Cit = 1111
Dr. Lic. = ND/ZEI201111
Lot No = 23411080A1
Cyl No = 04
Expiration Date = 10/01/2013
County = 10 Oper No. = 067518

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DAVID J. ZEIS

Remarks: