

ETHANOL BREATH STANDARD CYLINDER REPORT

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Ray Weber

Location: TRAIL CO. S.O.

Intoxilyzer Serial No.: 80-004200

Lot Number: 659338

Cylinder Number: #45

Expiration Date: 09/03/2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.081 %

Test 2: 0.080 %

Test 3: 0.080 %

(ATTACH TEST RECORD)

Ray Weber
Chemical Test Operator Signature

12/31/2009
Date

Deb Kashur
Reviewed By

25 Jan 2010
Date

80-004200
P. 1 of 2
YAN

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004200
Location = HILL 8164.13.00 06/09
12/31/2009 20:49

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	20:50
02 Std. Gas	0.081	20:50
03 Room Air	0.000	20:51
04 Std. Gas	0.080	20:51
05 Room Air	0.000	20:51
06 Std. Gas	0.080	20:52
07 Room Air	0.000	20:52

Lot No = 659358
Cyl No = 45
Exp Date = 09/03/2011
County = 49 Oper No. = 058405

Ray G. Weber

Operator Signature
RAY G. WEBER

Remarks: *ACA TEST*

Form 106-I8000

*80-004200
P. 2 of 2
JWH*