



# ETHANOL GAS STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <b>THOMAS HERZIG</b>		
Location <b>STIRMAN LEC (MAIN FLOOR)</b>	Intoxilyzer® Serial Number <b>80-004934</b>	
Gas Lot Number <b>19615080AS</b>	Gas Cylinder Number <b>35</b>	Gas Expiration Date <b>09/05/2017</b>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
  - A. 0.081 AC
  - B. 0.060 AC
  - C. 0.081 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature 	Date <b>12/31/15</b>
Reviewed By (Crime Laboratory Use Only) 	Date <b>11 Jan 16</b>

*main floor  
unit*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004934  
Location = STUT      8164.13.00 06/09  
12/31/2015      17:13

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:14
02 Std. Gas	0.081	17:14
03 Room Air	0.000	17:14
04 Std. Gas	0.080	17:15
05 Room Air	0.000	17:15
06 Std. Gas	0.081	17:16
07 Room Air	0.000	17:16

Lot No = 19615080A5  
Cyl No = 35  
Exp Date = 09/05/2017  
County = 47      Oper No. = 039706

  
Operator Signature  
THOMAS HERZIG

Remarks:

Form 106-I8000

LEC MAIN FLOOR  
NEW CYLINDER INSTALL