



# ETHANOL GAS STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>Thompson, Chad A #3524</i>		
Location <i>Cass Jail</i>	Intoxilyzer® Serial Number <i>80-006506</i>	
Gas Lot Number <i>01316080A2</i>	Gas Cylinder Number <i>33</i>	Gas Expiration Date <i>02/05/2018</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
  - A. *0.019* AC
  - B. *0.019* AC
  - C. *0.019* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i> <i>3524</i>	Date <i>09/18/2016</i> ✓
Reviewed By (Crime Laboratory Use Only) <i>Deb Shanaver</i>	Date <i>19 Sept 16</i>


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006506  
Location = CASS      8164.13.00 06/09  
09/18/2016      16:33

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:34
02 Std. Gas	0.079	16:34
03 Room Air	0.000	16:35
04 Std. Gas	0.079	16:35
05 Room Air	0.000	16:36
06 Std. Gas	0.079	16:36
07 Room Air	0.000	16:37

Lot No = 01316080A2  
Cyl No = 33  
Exp Date = 02/05/2018  
County = 09      Oper No. = 050910

 3524  
Operator Signature  
CHAD THOMPSON

Remarks: ACA Test for  
Cylinder Install  
Form 106-I8000