



ETHANOL GAS STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (10-2013)

Chemical Test Operator Name (Print) CHRIS ESTES		
Location DICKEY COUNTY SHERIFF DEPT/ ELLENDALE		Intoxilyzer® Serial Number 80-005953
Gas Lot Number 09216080A3	Gas Cylinder Number 43	Gas Expiration Date 05/05/2018

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. 0.083 AC
 - B. 0.081 AC
 - C. 0.082 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Chris Estes</i>	Date 11-03-16
Reviewed By (Crime Laboratory Use Only) <i>Deb Shanaver</i>	Date 03 NOV 16

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005953
Location = DCSO 8164.13.00 06/09
11/03/2016 09:02

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:03
02 Std. Gas	0.083	09:04
03 Room Air	0.000	09:04
04 Std. Gas	0.081	09:05
05 Room Air	0.000	09:05
06 Std. Gas	0.082	09:05
07 Room Air	0.000	09:06


Lot No = 09216080A3

Cyl No = 43

Exp Date = 05/05/2018

County = 11

Oper No. = 020504


Operator Signature
CHRIS ESTES

Remarks:

Form 106-I8000