

BREATH ALCOHOL EQUIPMENT
REPAIR AND MAINTENANCE

Instrument: 8000 Serial Number: 80-005951 Location: TOXL

1. Instrument repairs/maintenance performed:

- DIFFICULTIES CONNECTING TO INSTRUMENT AT WATF. BROUGHT BACK TO LAB TO CHECK MODEM CONNECTIONS. ALL ATTEMPTS TO CONNECT BY MODEM WERE SUCCESSFUL. FORMS LOADS, DOWNLOADS & FLASH UPDATE WERE ALL SUCCESSFUL.
- ISSUES WITH CONNECTION APPEAR TO BE PHONE LINE BETWEEN WATFORD CITY AND BISMARCK.
- REPLACED EXTERNAL BATTERY & TRAINED BATTERY

2. Does the instrument require further testing? Yes or No
If Yes, continue.
If No, sign and date below.

3. If needed, set the time, date, and location. N/A

4. Does the flow need to be calibrated? Yes or No
If Yes, attach paperwork.

5. Does the optical bench need to be calibrated? Yes or No
If Yes, attach test records.

6. Does the gas regulator need to be calibrated? Yes or No

PSI gauge reading _____ PSI display reading _____

Set instrument to Wet Bath and run tests:

7. Low AC. Use < 0.03 AC in ACA mode. Attach test record. _____

Sim SN: _____ Lot #: _____ AC: _____

8. Linearity Test. Use ≥ 0.25 AC in ACA mode. Attach test record. _____

Sim SN: _____ Lot #: _____ AC: _____

Set instrument to Gas and run tests:

9. Print test. Attach test record. _____

10. Interferent Check. Use a 0.10 AC ethanol plus 0.05% acetone in ABA mode.
Attach test record. _____

Sim SN: _____ Lot #: _____ AC: _____

11. RFI Check. Run in CMS mode. Key radio during any room air or subject test.
Attach test record. _____

12. Calibration Check. Use a valid 0.080 AC Ethanol Gas Standard. Run three
calibration sets in ACA mode. Attach test records. _____

Lot No. _____ Cylinder No. _____ Expiration Date _____


Test 1 _____	Test 1 _____	Test 1 _____
Test 2 _____	Test 2 _____	Test 2 _____
Test 3 _____	Test 3 _____	Test 3 _____

Average _____



Field Inspector's Signature

16 DEC. 16
Date



Reviewed by

20 Dec 16
Date