



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59281 (10-2016)

Serial Number <i>80-005956</i>	Instrument Location <i>Front Office - Oliver County Sheriff</i>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Properly Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. RFI Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if Any): _____
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer®8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>[Signature]</i>	Date <i>1-18-17</i>
---	------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified by <i>Deb Shanaver</i>	Certified Date <i>09 June 17</i>
--	-------------------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = OLIV 8164.14.00 09/16
01/18/2017 19:17

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 19615080A5
Standard Cyl #? 33
Standard Expiration? 09/05/2017
Oper No? 130683

Flow Cal. Date: 06/21/2013
Slope 706
Intercept -720901

IR Calibration Date: 04/02/2015
 3um 9um

0th Coef(*100): -23934 -22166
1st Coef(*100): 279293 133410
2nd Coef(*100): 3061 1559
H2O adj(mg/l*10k): 483 423

***** Printer Test End *****



Operator Signature
SHAINÉ TOWE

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = OLIV 8164.14.00 09/16
01/18/2017 19:18

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:19
02 Std. Gas	0.079	19:19
03 Room Air	0.000	19:19
04 Std. Gas	0.080	19:20
05 Room Air	0.000	19:20
06 Std. Gas	0.080	19:21
07 Room Air	0.000	19:21


Lot No = 19615080A5

Cyl No = 33

Exp Date = 09/05/2017

County = 33

Oper No. = 130683



Operator Signature
SHAINÉ TOWE

Remarks:

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = OLIV 8164.14.00 09/16
01/18/2017 19:32

Test	AC	Time
01 Room Air	0.000	19:33
02 *Subject Test	RFI*	19:34
03 Room Air	0.000	19:35

*Invalid Test
Inhibited - RFI

Sub Name = TEST, TEST TEST
Sub DOB = 01/18/2000
Sub Sex = Male Weight = 350
Test = OTH Cit = 0000000
Dr. Lic. = ND/0000000
Lot No = 19615080A5
Cyl No = 33
Expiration Date = 09/05/2017
County = 33 Oper No. = 130683



Operator Signature
SHAINE TOWE

Remarks:

Form 106-I8000