

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-004189

Location: TOXL

- Check When Done
- A. Is the breath tube heated? X
- B. Is the diagnostic check complete? X
- C. Is the time, date, and year correct (re-set if necessary)? X
- D. Print test. (Attach test record.) X
- E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN DR3846 Lot No. 16200 Rep AC 0.020 X
- F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR7111 Lot No. 16060 Rep AC 0.300 X
- G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.) X
Sim SN G10329 Lot No. 1CS4 Solution 0.10AC + 0.05% Acetone
- H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) X
- I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 17316080A2 Cyl No. 27 Exp. Date 5 Aug 18

Test 1 0.081 Test 4 0.080 Test 7 0.081

Test 2 0.080 Test 5 0.081 Test 8 0.080

Test 3 0.080 Test 6 0.080 Test 9 0.080

Average 0.080

J. Tank pressure, Level 3 Func D Sub G

⁶⁸⁰
P: ~~720~~ psi Reg 700 psi

K. Remarks/Maintenance record: Annual inspection; replaced o-ring on simulator return and breath hose.

18 Apr 17
Date

18 Apr 17
Date

Roberta Heuger-Timm
Field Inspector's Signature

Deb Shanaver
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
04/18/2017 13:26

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 17316080A2
Standard Cyl #? 27
Standard Expiration? 08/05/2018
Oper No? 888888

Flow Cal. Date: 02/02/2011
Slope 628
Intercept -452608

IR Calibration Date: 05/19/2016
 3um 9um

	3um	9um
0th Coef(*100):	-17767	-18761
1st Coef(*100):	263439	133606
2nd Coef(*100):	1960	1140
H2O adj(mg/l*10k):	341	362

***** Printer Test End *****



Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: *Print Test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
04/18/2017 13:29

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:30
02 Std. Sol.	0.020	13:30
03 Room Air	0.000	13:31
04 Std. Sol.	0.021	13:32
05 Room Air	0.000	13:32
06 Std. Sol.	0.021	13:33
07 Room Air	0.000	13:33

08 Sim Temp = 34.0°C

Simul Ser No = DR3846
Std Sol No = 16200
County = 08 Oper No. = 888888



Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: Low AC
0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
04/18/2017 13:35

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:35
02 Std. Sol.	0.305	13:36
03 Room Air	0.000	13:37
04 Std. Sol.	0.305	13:37
05 Room Air	0.000	13:38
06 Std. Sol.	0.306	13:39
07 Room Air	0.000	13:39

08 Sim Temp = 34.0°C

Simul Ser No = DR7111
Std Sol No = 16060
County = 08 Oper No. = 888888


Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: High AC
 0.300 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
04/18/2017 13:40

Test	AC	Time
01 Room Air	0.000	13:41
02 *Subject Test	INT*	13:41
03 Room Air	0.000	13:42

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS821456
Lot No = 17316080A2
Cyl No = 27
Expiration Date = 08/05/2018
County = 08 Oper No. = 888888

Roberta Grieger-Nimmo
Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: *Interferent Check*
0.10 AC + 0.05% Acetone

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
04/18/2017 13:43

Test	AC	Time
01 Diagnostic	OK	13:44
02 Room Air	0.000	13:44
03 Subject Test 1	0.000	13:45
04 Room Air	RFI*	13:46
05 Room Air	0.000	13:46

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS821456
Lot No = 17316080A2
Cyl No = 27
Expiration Date = 08/05/2018
County = 08 Oper No. = 888888

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.


Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: RFI Check


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
04/18/2017 13:50

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:51
02 Std. Gas	0.081	13:51
03 Room Air	0.000	13:52
04 Std. Gas	0.080	13:52
05 Room Air	0.000	13:52
06 Std. Gas	0.080	13:53
07 Room Air	0.000	13:53

Lot No = 17316080A2
Cyl No = 27
Exp Date = 08/05/2018
County = 08 Oper No. = 888888


Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: Calibration Check
0.080 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
04/18/2017 13:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:55
02 Std. Gas	0.080	13:55
03 Room Air	0.000	13:55
04 Std. Gas	0.081	13:56
05 Room Air	0.000	13:56
06 Std. Gas	0.080	13:57
07 Room Air	0.000	13:57

Lot No = 17316080A2
Cyl No = 27
Exp Date = 08/05/2018
County = 08 Oper No. = 888888


Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: Calibration Check
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = TOXL 8164.14.00 09/16
04/18/2017 13:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:58
02 Std. Gas	0.081	13:59
03 Room Air	0.000	13:59
04 Std. Gas	0.080	14:00
05 Room Air	0.000	14:00
06 Std. Gas	0.080	14:00
07 Room Air	0.000	14:01

Lot No = 17316080A2
Cyl No = 27
Exp Date = 08/05/2018
County = 08 Oper No. = 888888



Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: Calibration Check
0.080 AC

Form 106-I8000