



ETHANOL GAS STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (10-2013)

Chemical Test Operator Name (Print) HUNTER RHOADES / FLD INSP.		
Location MANNING, ND - DAWN COUNTY S.O.	Intoxilyzer® Serial Number 80-005155 80-005955 Deb Shenaver	
Gas Lot Number 19615080A5	Gas Cylinder Number 24	Gas Expiration Date 9/5/2017

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. 0. 083 AC
 - B. 0. 083 AC
 - C. 0. 083 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Hunter Rhoades</i>	Date 4/28/17
Reviewed By (Crime Laboratory Use Only) <i>Deb Shenaver</i>	Date 03 May 17


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005955
Location = DUNN 8164.14.00 09/16
04/28/2017 08:57

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:58
02 Std. Gas	0.083	08:58
03 Room Air	0.000	08:59
04 Std. Gas	0.083	08:59
05 Room Air	0.000	09:00
06 Std. Gas	0.083	09:00
07 Room Air	0.000	09:01

Lot No = 19615080A5
Cyl No = 24
Exp Date = 09/05/2017
County = 13 Oper No. = 130635



Operator Signature
HUNTER RHOADES

Remarks: