

CHAIN OF CUSTODY
BREATH ALCOHOL INSTRUMENT

Instrument: 8000 Serial Number: 80-006494

Agency being assigned instrument: Crime Lab

Date 08 May 17

Transferred from: Valley City PD

Transferred to:

Name (print) Deb Shanaver

Name (print) Deb Shanaver

Signature Deb Shanaver

Signature Deb Shanaver

Agency Crime Lab

Agency Crime Lab

Reason for transfer/remarks:

Falling out of calibration.
High AC is reading too high.

Or Shipped Via:

USPS UPS FedEx Other

Date Sending Instrument: / / Air Bill No.