

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-003067

Location: TOXL

- |   | Check When Done |
|---|-----------------|
| A. Is the breath tube heated?   | <u>X</u>        |
| B. Is the diagnostic check complete?  | <u>X</u>        |
| C. Is the time, date, and year correct (re-set if necessary)?   | <u>X</u>        |
| D. Print test. (Attach test record.)  | <u>X</u>        |
| E. Low AC. Use $\leq 0.03$ AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR3846</u> Lot No. <u>16200</u> Rep AC <u>0.020 AC</u>  | <u>X</u>        |
| F. High AC. Use $\geq 0.25$ AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR7111</u> Lot No. <u>16060</u> Rep AC <u>0.300 AC</u>   | <u>X</u>        |
| G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.<br>(Attach test record.)<br>Sim SN <u>DR5135</u> Lot No. <u>1CS4</u> Solution <u>0.10 AC + 0.05%</u><br><span style="margin-left: 150px;">Acetone</span> | <u>X</u>        |
| H. RFI check. Run CMS mode. Key radio during CMS test.<br>(Attach test record.)   | <u>X</u>        |
| I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.<br>(Attach test record.) <u>0.080</u> AC   |                 |

Lot No. 17316080A2 Cyl No. 27 Exp. Date 5 Aug 18  
Test 1 0.080 Test 4 0.081 Test 7 0.080  
Test 2 0.080 Test 5 0.080 Test 8 0.081  
Test 3 0.080 Test 6 0.080 Test 9 0.081  
Average 0.080

J. Tank pressure, Level 3 Func D Sub G P: 388 psi Reg 400 psi

K. Remarks/Maintenance record: Annual inspection; Changed O-ring on Simulator return and breath hose.

9 May 17  
Date  
5/9/2017  
Date

Roberto Huenegui-Himeno  
Field Inspector's Signature  
Charles E. Eddy  
Reviewed by

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-003067  
Location = TOXL                    8164.14.00 09/16  
05/08/2017                            13:27

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            17316080A2  
Standard Cyl #?            33  
Standard Expiration?        08/05/2018  
Oper No?                    888888

Flow Cal. Date:            01/23/2013  
Slope                        664  
Intercept                   -509770

IR Calibration Date:        04/30/2015  
                                  3um                    9um

-----  
0th Coef(\*100):            -19633                -12115  
1st Coef(\*100):            278224                137001  
2nd Coef(\*100):            2596                   1415  
H2O adj(mg/l\*10k):        323                    247

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: Print test

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003067  
Location = TOXL                    8164.14.00 09/16  
05/09/2017                            10:04

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:05
02 Std. Sol.	0.021	10:06
03 Room Air	0.000	10:06
04 Std. Sol.	0.020	10:07
05 Room Air	0.000	10:07
06 Std. Sol.	0.020	10:08
07 Room Air	0.000	10:09

08 Sim Temp = 34.0°C

Simul Ser No = DR3846  
Std Sol No = 16200  
County = 08                            Oper No. = 888888

  
Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: low AC  
0.020

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003067  
Location = TOXL      8164.14.00 09/16  
05/09/2017      10:09

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:10
02 Std. Sol.	0.301	10:11
03 Room Air	0.000	10:11
04 Std. Sol.	0.303	10:12
05 Room Air	0.000	10:13
06 Std. Sol.	0.302	10:13
07 Room Air	0.000	10:14

08 Sim Temp = 34.0°C

Simul Ser No = DR7111  
Std Sol No = 16060  
County = 08      Oper No. = 888888

  
Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: High AC  
0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-003067  
Location = TOXL                    8164.14.00 09/16  
05/08/2017                            13:47

Test	AC	Time
01 Room Air	0.000	13:47
02 *Subject Test	INT*	13:48
03 Room Air	0.000	13:49

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = INTERFERENT  
Dr. Lic. = ND/DIS821456  
Lot No = 17316080A2  
Cyl No = 27  
Expiration Date = 08/05/2018  
County = 08                            Oper No. = 888888



Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: Interferent Check  
0.10AC + 0.05% Acetone

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-003067  
Location = TOXL                    8164.14.00 09/16  
05/08/2017                            13:42

Test	AC	Time
01 Diagnostic	OK	13:43
02 Room Air	0.000	13:43
03 *Subject Test	RFI*	13:43
04 Room Air	0.000	13:44

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI CHECK  
Dr. Lic. = ND/DIS821456  
Lot No = 17316080A2  
Cyl No = 27  
Expiration Date = 08/05/2018  
County = 08                            Oper No. = 888888

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks:

RFI Check


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003067  
Location = TOXL      8164.14.00 09/16  
05/08/2017      13:31

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:32
02 Std. Gas	0.080	13:32
03 Room Air	0.000	13:33
04 Std. Gas	0.080	13:33
05 Room Air	0.000	13:34
06 Std. Gas	0.080	13:34
07 Room Air	0.000	13:35

Lot No = 17316080A2  
Cyl No = 27  
Exp Date = 08/05/2018  
County = 08      Oper No. = 888888

  
Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: Calibration Check  
0.080 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003067  
Location = TOXL      8164.14.00 09/16  
05/08/2017      13:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:35
02 Std. Gas	0.081	13:36
03 Room Air	0.000	13:36
04 Std. Gas	0.080	13:37
05 Room Air	0.000	13:37
06 Std. Gas	0.080	13:37
07 Room Air	0.000	13:38

Lot No = 17316080A2  
Cyl No = 27  
Exp Date = 08/05/2018  
County = 08      Oper No. = 888888

  
Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: Calibration Check  
0.080 AC




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003067  
Location = TOXL                8164.14.00 09/16  
05/08/2017                        13:38

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:39
02 Std. Gas	0.080	13:39
03 Room Air	0.000	13:40
04 Std. Gas	0.081	13:40
05 Room Air	0.000	13:41
06 Std. Gas	0.081	13:41
07 Room Air	0.000	13:41

Lot No = 17316080A2  
Cyl No = 27  
Exp Date = 08/05/2018  
County = 08                      Oper No. = 888888

  
Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: Calibration Check  
          0.080 AC