

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-005947

Location: STARK COUNTY L.E.C. (DKPD)

- Check When Done
- A. Is the breath tube heated? ✓
 - B. Is the diagnostic check complete? ✓
 - C. Is the time, date, and year correct (re-set if necessary)? ✓
 - D. Print test. (Attach test record.) ✓
 - E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5113 Lot No. 201507F Rep AC 0.020 ✓
 - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3378 Lot No. 16060 Rep AC 0.300 ✓
 - G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN DR3803 Lot No. ICS4 Solution 0.10AC + 0.05% ACETONE ✓
 - H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) ✓
 - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 09216080A3 Cyl No. 45 Exp. Date 5/5/2018

Test 1 <u>0.079</u>	Test 4 <u>0.079</u>	Test 7 <u>0.079</u>
Test 2 <u>0.078</u>	Test 5 <u>0.079</u>	Test 8 <u>0.079</u>
Test 3 <u>0.079</u>	Test 6 <u>0.079</u>	Test 9 <u>0.079</u>

Average 0.079

J. Tank pressure, Level 3 Func D Sub G P: 708 psi Reg 700 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION, ADJUSTED

GAS REGULATOR
5/11/2017
Date

Charles E. Ed
Field Inspector's Signature

12 May 17
Date

Roberta Huggins-Kemmo
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005947
Location = DKPD 8164.14.00 09/16
05/11/2017 13:26

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 09216080A3
Standard Cyl #? 45
Standard Expiration? 05/05/2018
Oper No? 666666

Flow Cal. Date: 06/21/2013
Slope 693
Intercept -735360

IR Calibration Date: 04/06/2015
 3um 9um

0th Coef(*100): -28096 -26982
1st Coef(*100): 280034 136042
2nd Coef(*100): 3530 1567
H2O adj(mg/l*10k): 510 437

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005947
Location = DKPD 8164.14.00 09/16
05/11/2017 13:38

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:39
02 Std. Sol.	0.017	13:40
03 Room Air	0.000	13:40
04 Std. Sol.	0.017	13:41
05 Room Air	0.000	13:42
06 Std. Sol.	0.017	13:42
07 Room Air	0.000	13:43

08 Sim Temp = 34.0°C

Simul Ser No = DR5113

Std Sol No = 201507F

County = 45

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

LOW AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005947
Location = DKPD 8164.14.00 09/16
05/11/2017 13:33

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:34
02 Std. Sol.	0.300	13:35
03 Room Air	0.000	13:35
04 Std. Sol.	0.303	13:36
05 Room Air	0.000	13:36
06 Std. Sol.	0.303	13:37
07 Room Air	0.000	13:38

08 Sim Temp = 34.0°C

Simul Ser No = DR3378

Std Sol No = 16060

County = 45

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005947
Location = DKPD 8164.14.00 09/16
05/11/2017 13:29

Test	AC	Time
01 Room Air	0.000	13:30
02 *Subject Test	INT*	13:31
03 Room Air	0.000	13:31

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 09216080A3
Cyl No = 45
Expiration Date = 05/05/2018
County = 45 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent
0.10AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005947
Location = DKPD 8164.14.00 09/16
05/11/2017 13:27

Test	AC	Time
01 Diagnostic	OK	13:27
02 Room Air	0.000	13:28
03 *Subject Test	RFI*	13:28
04 Room Air	0.000	13:29

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 09216080A3
Cyl No = 45
Expiration Date = 05/05/2018
County = 45 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005947
Location = DKPD 8164.14.00 09/16
05/11/2017 13:43

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:44
02 Std. Gas	0.079	13:44
03 Room Air	0.000	13:45
04 Std. Gas	0.078	13:45
05 Room Air	0.000	13:46
06 Std. Gas	0.079	13:46
07 Room Air	0.000	13:47

Lot No = 09216080A3
Cyl No = 45
Exp Date = 05/05/2018
County = 45 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005947
Location = DKPD 8164.14.00 09/16
05/11/2017 13:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:47
02 Std. Gas	0.079	13:48
03 Room Air	0.000	13:48
04 Std. Gas	0.079	13:49
05 Room Air	0.000	13:49
06 Std. Gas	0.079	13:49
07 Room Air	0.000	13:50

Lot No = 09216080A3

Cyl No = 45

Exp Date = 05/05/2018

County = 45

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005947
Location = DKPD 8164.14.00 09/16
05/11/2017 13:50

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:51
02 Std. Gas	0.079	13:51
03 Room Air	0.000	13:52
04 Std. Gas	0.079	13:52
05 Room Air	0.000	13:53
06 Std. Gas	0.079	13:53
07 Room Air	0.000	13:53

Lot No = 09216080A3
Cyl No = 45
Exp Date = 05/05/2018
County = 45 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000