

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-005955

Location: DUNN COUNTY S.O. (DUNN)

Check When Done

- A. Is the breath tube heated?
- B. Is the diagnostic check complete? ✓
- C. Is the time, date, and year correct (re-set if necessary)? ✓
- D. Print test. (Attach test record.) ✓
- E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
 Sim SN DR5113 Lot No. 201507F Rep AC 0.020 ✓
- F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
 Sim SN DR3378 Lot No. 16060 Rep AC 0.300 ✓
- G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
 (Attach test record.)
 Sim SN DR3803 Lot No. ICS4 Solution 0.10AC+0.05% ACETONE ✓
- H. RFI check. Run CMS mode. Key radio during CMS test.
 (Attach test record.) ✓
- I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
 (Attach test record.) 0.080 AC

Lot No. 19615080A5 Cyl No. 24 Exp. Date 9/5/2017
 Test 1 0.081 Test 4 0.082 Test 7 0.082
 Test 2 0.082 Test 5 0.083 Test 8 0.092
 Test 3 0.082 Test 6 0.081 Test 9 0.081
 Average 0.082

J. Tank pressure, Level 3 Func D Sub G P: 1068 psi Reg 1000 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION

5/11/2017
 Date

Charles E. Ed
 Field Inspector's Signature

12 May 17
 Date

Roberta Huggins
 Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005955
Location = DUNN 8164.14.00 09/16
05/11/2017 12:09

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNPNPQRSTUVWXYZ!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNPNPQRSTUVWXYZ!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 19615080A5
Standard Cyl #? 24
Standard Expiration? 09/05/2017
Oper No? 666666

Flow Cal. Date: 06/21/2013
Slope 757
Intercept -837385

IR Calibration Date: 06/21/2013
3um 9um

0th Coef(*100):	-19897	-22878
1st Coef(*100):	268381	131979
2nd Coef(*100):	5157	1473
H2O adj(mg/l*10k):	717	600

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005955
Location = DUNN 8164.14.00 09/16
05/11/2017 12:34

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:35
02 Std. Sol.	0.018	12:36
03 Room Air	0.000	12:36
04 Std. Sol.	0.019	12:37
05 Room Air	0.000	12:37
06 Std. Sol.	0.019	12:38
07 Room Air	0.000	12:39

08 Sim Temp = 34.0°C

Simul Ser No = DR5113
Std Sol No = 201507F
County = 13 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Low AC*
0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005955
Location = DUNN 8164.14.00 09/16
05/11/2017 12:29

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:30
02 Std. Sol.	0.300	12:31
03 Room Air	0.000	12:31
04 Std. Sol.	0.302	12:32
05 Room Air	0.000	12:32
06 Std. Sol.	0.304	12:33
07 Room Air	0.000	12:34

08 Sim Temp = 34.0°C

Simul Ser No = DR3378
Std Sol No = 16060
County = 13 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *HIGH AC*
 0.300 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005955
Location = DUNN 8164.14.00 09/16
05/11/2017 12:13

Test	AC	Time
01 Room Air	0.000	12:14
02 *Subject Test	INT*	12:14
03 Room Air	0.000	12:15

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 19615080A5
Cyl No = 24
Expiration Date = 09/05/2017
County = 13 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent
0.10ACT + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005955
Location = DUNN 8164.14.00 09/16
05/11/2017 12:10

Test	AC	Time
01 Diagnostic	OK	12:11
02 Room Air	0.000	12:12
03 *Subject Test	RFI*	12:12
04 Room Air	0.000	12:12

*Invalid Test
 Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 19615080A5
Cyl No = 24
Expiration Date = 09/05/2017
County = 13 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *RFI CHECK*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005955
Location = DUNN 8164.14.00 09/16
05/11/2017 12:16

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:17
02 Std. Gas	0.081	12:17
03 Room Air	0.000	12:18
04 Std. Gas	0.082	12:18
05 Room Air	0.000	12:19
06 Std. Gas	0.082	12:19
07 Room Air	0.000	12:20

Lot No = 19615080A5
Cyl No = 24
Exp Date = 09/05/2017
County = 13 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005955
Location = DUNN 8164.14.00 09/16
05/11/2017 12:20

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:20
02 Std. Gas	0.082	12:21
03 Room Air	0.000	12:21
04 Std. Gas	0.083	12:21
05 Room Air	0.000	12:22
06 Std. Gas	0.081	12:22
07 Room Air	0.000	12:23

Lot No = 19615080A5
Cyl No = 24
Exp Date = 09/05/2017
County = 13 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005955
Location = DUNN 8164.14.00 09/16
05/11/2017 12:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:40
02 Std. Gas	0.082	12:40
03 Room Air	0.000	12:41
04 Std. Gas	0.082	12:41
05 Room Air	0.000	12:42
06 Std. Gas	0.081	12:42
07 Room Air	0.000	12:43

Lot No = 19615080A5
Cyl No = 24
Exp Date = 09/05/2017
County = 13 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC