

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-006500 Location: HARVEY P.D (HARV)

- | | Check When Done |
|--|-----------------|
| A. Is the breath tube heated? | <u>✓</u> |
| B. Is the diagnostic check complete? | <u>✓</u> |
| C. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| D. Print test. (Attach test record.) | <u>✓</u> |
| E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR5113</u> Lot No. <u>201507E</u> Rep AC <u>0.020</u> | <u>✓</u> |
| F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR337B</u> Lot No. <u>16060</u> Rep AC <u>0.300</u> | <u>✓</u> |
| G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN <u>DR3803</u> Lot No. <u>ICS4</u> Solution <u>0.10AC + 0.05% ACETONE</u> | <u>✓</u> |
| H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) | <u>✓</u> |
| I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) <u>0.080</u> AC | |

Lot No. 19615080A5 Cyl No. 36 Exp. Date 9/5/2017

Test 1 <u>0.080</u>	Test 4 <u>0.080</u>	Test 7 <u>0.081</u>
Test 2 <u>0.080</u>	Test 5 <u>0.080</u>	Test 8 <u>0.081</u>
Test 3 <u>0.080</u>	Test 6 <u>0.080</u>	Test 9 <u>0.080</u>

Average 0.080

J. Tank pressure, Level 3 Func D Sub G P: 956 psi Reg 950 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION

5/12/2017
Date

12 May 17
Date

Charles E. Ed
Field Inspector's Signature

Deb Shanever
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = HARV 8164.13.00 06/09
05/12/2017 09:30

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-==|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-==|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 19615080A5
Standard Cyl #? 36
Standard Expiration? 09/05/2017
Oper No? 666666

Flow Cal. Date: 08/18/2015
Slope 681
Intercept -680985

IR Calibration Date: 08/18/2015
 3um 9um

	3um	9um
0th Coef(*100):	-3148	-17259
1st Coef(*100):	254475	138520
2nd Coef(*100):	1954	1329
H2O adj(mg/l*10k):	193	414

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = HARV 8164.13.00 06/09
05/12/2017 09:57

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:57
02 Std. Sol.	0.020	09:58
03 Room Air	0.000	09:59
04 Std. Sol.	0.020	09:59
05 Room Air	0.000	10:00
06 Std. Sol.	0.020	10:01
07 Room Air	0.000	10:01

08 Sim Temp = 34.0°C

Simul Ser No = DR5113

Std Sol No = 201507F

County = 52

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

LOW AC

0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = HARV 8164.13.00 06/09
05/12/2017 09:50

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:51
02 Std. Sol.	0.310	09:52
03 Room Air	0.000	09:52
04 Std. Sol.	0.310	09:53
05 Room Air	0.000	09:54
06 Std. Sol.	0.310	09:54
07 Room Air	0.000	09:55

08 Sim Temp = 34.0°C

Simul Ser No = DR3378

Std Sol No = 16060

County = 52

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = HARV 8164.13.00 06/09
05/12/2017 09:37

Test	AC	Time
01 Room Air	0.000	09:38
02 *Subject Test	INT*	09:38
03 Room Air	0.000	09:39

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 19615080A5
Cyl No = 36
Expiration Date = 09/05/2017
County = 52 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Interferent*
0.10AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = HARV 8164.13.00 06/09
05/12/2017 09:39

Test	AC	Time
01 Diagnostic	OK	09:40
02 Room Air	0.000	09:41
03 *Subject Test	RFI*	09:41
04 Room Air	0.000	09:41

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 19615080A5
Cyl No = 36
Expiration Date = 09/05/2017
County = 52 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = HARV 8164.13.00 06/09
05/12/2017 09:33

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:34
02 Std. Gas	0.080	09:34
03 Room Air	0.000	09:35
04 Std. Gas	0.080	09:35
05 Room Air	0.000	09:36
06 Std. Gas	0.080	09:36
07 Room Air	0.000	09:37

Lot No = 19615080A5
Cyl No = 36
Exp Date = 09/05/2017
County = 52 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: ~~CEX~~ CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = HARV 8164.13.00 06/09
05/12/2017 09:41

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:42
02 Std. Gas	0.080	09:42
03 Room Air	0.000	09:43
04 Std. Gas	0.080	09:43
05 Room Air	0.000	09:44
06 Std. Gas	0.080	09:44
07 Room Air	0.000	09:45

Lot No = 19615080A5
Cyl No = 36
Exp Date = 09/05/2017
County = 52 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

