

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-005953

Location: DCSO
Dickey Co. SO

Check When Done

- A. Is the breath tube heated? X
- B. Is the diagnostic check complete? X
- C. Is the time, date, and year correct (re-set if necessary)? X
- D. Print test. (Attach test record.) X
- E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5113 Lot No. 201507F Rep AC 0.020 AC X
- F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3378 Lot No. 16060 Rep AC 0.300 AC X
- G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN DR3803 Lot No. 1CS4 Solution 0.10 AC + 0.05% Acetone X
- H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) X

I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 092160780A3 Cyl No. 43 Exp. Date 5 May 18

Test 1 0.081 Test 4 0.081 Test 7 0.081

Test 2 0.081 Test 5 0.081 Test 8 0.081

Test 3 0.081 Test 6 0.080 Test 9 0.082

Average 0.081

J. Tank pressure, Level 3 Func D Sub G P: 1000 psi Reg 1000 psi

K. Remarks/Maintenance record: Annual inspection; Changed O-ring on
breath hose

15 May 17
Date

Roberta Huggs-Rimmo
Field Inspector's Signature

15 May 17
Date

Deb Shanaver
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005953
Location = DCSO 8164.13.00 06/09
05/15/2017 15:26

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 09216080A3
Standard Cyl #? 43
Standard Expiration? 05/05/2018
Oper No? 888888

Flow Cal. Date: 06/16/2015
Slope 678
Intercept -668987

IR Calibration Date: 04/30/2015
3um 9um

0th Coef(*100): -13435 -19950
1st Coef(*100): 257165 129445
2nd Coef(*100): 3995 1631
H2O adj(mg/l*10k): 505 548

***** Printer Test End *****



Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: Print Test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005953
Location = DCSO 8164.13.00 06/09
05/15/2017 15:44

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:44
02 Std. Sol.	0.019	15:45
03 Room Air	0.000	15:46
04 Std. Sol.	0.018	15:46
05 Room Air	0.000	15:47
06 Std. Sol.	0.018	15:48
07 Room Air	0.000	15:48


08 Sim Temp = 34.0°C

Simul Ser No = DR5113

Std Sol No = 201507F

County = 11

Oper No. = 888888



Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks:

Low AC

~~0.080 AC RW~~

0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005953
Location = DCSO 8164.13.00 06/09
05/15/2017 15:55

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:56
02 Std. Sol.	0.306	15:57
03 Room Air	0.000	15:57
04 Std. Sol.	0.305	15:58
05 Room Air	0.000	15:59
06 Std. Sol.	0.305	15:59
07 Room Air	0.000	16:00


08 Sim Temp = 34.0°C

Simul Ser No = DR3378

Std Sol No = 16060

County = 11

Oper No. = 888888



Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: High AC
0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005953
Location = DCSO 8164.13.00 06/09
05/15/2017 16:02

Test	AC	Time
01 Room Air	0.000	16:02
02 *Subject Test	INT*	16:03
03 Room Air	0.000	16:04

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 09216080A3
Cyl No = 43
Expiration Date = 05/05/2018
County = 11 Oper No. = 888888



Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: Interferent Check
0.100 AC + 0.05% Acetone

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005953
Location = DCSO 8164.13.00 06/09
05/15/2017 15:37

Test	AC	Time
01 Diagnostic	OK	15:38
02 Room Air	0.000	15:39
03 *Subject Test	RFI*	15:39
04 Room Air	0.000	15:39

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI
Dr. Lic. = ND/DIS821456
Lot No = 09216080A3
Cyl No = 43
Expiration Date = 05/05/2018
County = 11 Oper No. = 888888

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Roberta Grieger-Nimmo

Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: RFI Check

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005953
Location = DCSO 8164.13.00 06/09
05/15/2017 15:26

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:27
02 Std. Gas	0.081	15:27
03 Room Air	0.000	15:28
04 Std. Gas	0.081	15:28
05 Room Air	0.000	15:29
06 Std. Gas	0.081	15:29
07 Room Air	0.000	15:30


Lot No = 09216080A3

Cyl No = 43

Exp Date = 05/05/2018

County = 11

Oper No. = 888888



Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: Calibration Check

0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005953
Location = DCSO 8164.13.00 06/09
05/15/2017 15:30

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:30
02 Std. Gas	0.081	15:31
03 Room Air	0.000	15:31
04 Std. Gas	0.081	15:32
05 Room Air	0.000	15:32
06 Std. Gas	0.080	15:32
07 Room Air	0.000	15:33

Lot No = 09216080A3
Cyl No = 43
Exp Date = 05/05/2018
County = 11 Oper No. = 888888

Roberta Grieger-Nimmo
Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: Calibration Check
0.080 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005953
Location = DCSO 8164.13.00 06/09
05/15/2017 15:33

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:34
02 Std. Gas	0.081	15:34
03 Room Air	0.000	15:35
04 Std. Gas	0.081	15:35
05 Room Air	0.000	15:36
06 Std. Gas	0.082	15:36
07 Room Air	0.000	15:36

Lot No = 09216080A3
Cyl No = 43
Exp Date = 05/05/2018
County = 11 Oper No. = 888888


Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: Calibration Check
0.080 AC
Form 106-I8000