

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-005957

Location: CROS (DIVIDE COUNTY S.O.)

Check When Done

- A. Is the breath tube heated?
- B. Is the diagnostic check complete?
- C. Is the time, date, and year correct (re-set if necessary)?
- D. Print test. (Attach test record.)
- E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5113 Lot No. 201507F Rep AC 0.020
- F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3378 Lot No. 16060 Rep AC 0.300
- G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN PR3803 Lot No. ICS4 Solution 0.10 AC + 0.05% ACETONE
- H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.)
- I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 19615080A5 Cyl No. 23 Exp. Date 9/5/2017

Test 1 <u>0.079</u>	Test 4 <u>0.078</u>	Test 7 <u>0.079</u>
Test 2 <u>0.079</u>	Test 5 <u>0.079</u>	Test 8 <u>0.078</u>
Test 3 <u>0.079</u>	Test 6 <u>0.079</u>	Test 9 <u>0.079</u>

Average 0.079

J. Tank pressure, Level 3 Func D Sub G P: 994 psi Reg 1000 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION

5/18/2017
Date

Charles E. Egan
Field Inspector's Signature

19 May 17
Date

Deb Shanaver
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005957
Location = CROS 8164.14.00 09/16
05/18/2017 09:07

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 19615080A5
Standard Cyl #? 23
Standard Expiration? 09/05/2017
Oper No? 666666

Flow Cal. Date: 06/21/2013
Slope 647
Intercept -613317

IR Calibration Date: 05/11/2015
 3um 9um

0th Coef(*100): -16666 -18045
1st Coef(*100): 278478 132949
2nd Coef(*100): 3303 1397
H2O adj(mg/l*10k): 379 351

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks: PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005957
Location = CROS 8164.14.00 09/16
05/18/2017 09:51

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:51
02 Std. Sol.	0.019	09:52
03 Room Air	0.000	09:52
04 Std. Sol.	0.020	09:53
05 Room Air	0.000	09:54
06 Std. Sol.	0.019	09:54
07 Room Air	0.000	09:55

08 Sim Temp = 34.0°C

DR5113

Simul Ser No = ~~DR3378~~ **CEE**

Std Sol No = ~~16060~~ **CEE**

County = 12 **201507F** Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005957
Location = CROS 8164.14.00 09/16
05/18/2017 09:44

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:45
02 Std. Sol.	0.310	09:46
03 Room Air	0.000	09:46
04 Std. Sol.	0.310	09:47
05 Room Air	0.000	09:48
06 Std. Sol.	0.310	09:48
07 Room Air	0.000	09:49


08 Sim Temp = 34.0°C

Simul Ser No = DR3378

Std Sol No = 16060

County = 12

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005957
Location = CROS 8164.14.00 09/16
05/18/2017 09:35

Test	AC	Time
01 Room Air	0.000	09:36
02 *Subject Test	INT*	09:36
03 Room Air	0.000	09:37

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 19615080A5
Cyl No = 23
Expiration Date = 09/05/2017
County = 12 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Interferent*
0.10 AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005957
Location = CROS 8164.14.00 09/16
05/18/2017 09:33

Test	AC	Time
01 Diagnostic	OK	09:34
02 Room Air	0.000	09:34
03 *Subject Test	RFI*	09:34
04 Room Air	0.000	09:35

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 19615080A5
Cyl No = 23
Expiration Date = 09/05/2017
County = 12 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005957
Location = CROS 8164.14.00 09/16
05/18/2017 09:24

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:25
02 Std. Gas	0.079	09:25
03 Room Air	0.000	09:26
04 Std. Gas	0.079	09:26
05 Room Air	0.000	09:27
06 Std. Gas	0.079	09:27
07 Room Air	0.000	09:27

Lot No = 19615080A5
Cyl No = 23
Exp Date = 09/05/2017
County = 12 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005957
Location = CROS 8164.14.00 09/16
05/18/2017 09:28

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:29
02 Std. Gas	0.078	09:29
03 Room Air	0.000	09:29
04 Std. Gas	0.079	09:30
05 Room Air	0.000	09:30
06 Std. Gas	0.079	09:31
07 Room Air	0.000	09:31

Lot No = 19615080A5
Cyl No = 23
Exp Date = 09/05/2017
County = 12 Oper No. = 666666


Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005957
Location = CROS 8164.14.00 09/16
05/18/2017 09:56

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:57
02 Std. Gas	0.079	09:57
03 Room Air	0.000	09:58
04 Std. Gas	0.078	09:58
05 Room Air	0.000	09:59
06 Std. Gas	0.079	09:59
07 Room Air	0.000	10:00

Lot No = 19615080A5
Cyl No = 23
Exp Date = 09/05/2017
County = 12 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080AC