

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-004942

Location: BURKE COUNTY S.O. (BURK)

- | | Check When Done |
|--|-----------------|
| A. Is the breath tube heated? | <u>✓</u> |
| B. Is the diagnostic check complete? | <u>✓</u> |
| C. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| D. Print test. (Attach test record.) | <u>✓</u> |
| E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR5113</u> Lot No. <u>201507F</u> Rep AC <u>0.020</u> | <u>✓</u> |
| F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3378</u> Lot No. <u>16060</u> Rep AC <u>0.300</u> | <u>✓</u> |
| G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN <u>DR3803</u> Lot No. <u>ICS4</u> Solution <u>0.10AC + 0.05% ACETONE</u> | <u>✓</u> |
| H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) | <u>✓</u> |
| I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) <u>0.080</u> AC | |

Lot No. <u>19615080A5</u>	Cyl No. <u>018</u>	Exp. Date <u>9/5/2017</u>
Test 1 <u>0.082</u>	Test 4 <u>0.081</u>	Test 7 <u>0.083</u>
Test 2 <u>0.082</u>	Test 5 <u>0.082</u>	Test 8 <u>0.083</u>
Test 3 <u>0.082</u>	Test 6 <u>0.082</u>	Test 9 <u>0.082</u>
Average <u>0.082</u>		

J. Tank pressure, Level 3 Func D Sub G P: 1005 psi Reg: 000 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION

5/17/2017
Date

Charles E. Edr
Field Inspector's Signature

22 May 17
Date

Deb Shanaver
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = BURK 8164.13.00 06/09
05/17/2017 11:16

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 19615080A5
Standard Cyl #? 18
Standard Expiration? 09/05/2017
Oper No? 666666

Flow Cal. Date: 08/21/2012
Slope 638
Intercept -487461

IR Calibration Date: 05/12/2015
 3um 9um

	3um	9um
0th Coef(*100):	-11858	-18295
1st Coef(*100):	255387	131150
2nd Coef(*100):	3253	1510
H2O adj(mg/l*10k):	493	483

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks: PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = BURK 8164.13.00 06/09
05/17/2017 11:36

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:36
02 Std. Sol.	0.017	11:37
03 Room Air	0.000	11:38
04 Std. Sol.	0.018	11:38
05 Room Air	0.000	11:39
06 Std. Sol.	0.018	11:40
07 Room Air	0.000	11:40

08 Sim Temp = 34.0°C

Simul Ser No = DR5113

Std Sol No = 201507F

County = 07

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

Low AC

0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = BURK 8164.13.00 06/09
05/17/2017 11:30

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:30
02 Std. Sol.	0.296	11:31
03 Room Air	0.000	11:32
04 Std. Sol.	0.297	11:32
05 Room Air	0.000	11:33
06 Std. Sol.	0.299	11:34
07 Room Air	0.000	11:34

08 Sim Temp = 34.0°C

Simul Ser No = DR3378
Std Sol No = 16060
County = 07 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: HIGH AC
 0.300 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = BURK 8164.13.00 06/09
05/17/2017 11:42

Test	AC	Time
01 Room Air	0.000	11:43
02 *Subject Test	INT*	11:43
03 Room Air	0.000	11:44

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 19615080A5
Cyl No = 18
Expiration Date = 09/05/2017
County = 07 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: ^{CEE} ~~Int~~ INTERFERENT

0.10 AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = BURK 8164.13.00 06/09
05/17/2017 11:44

Test	AC	Time
01 Diagnostic	OK	11:45
02 Room Air	0.000	11:45
03 *Subject Test	RFI*	11:46
04 Room Air	0.000	11:46

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 19615080A5
Cyl No = 18
Expiration Date = 09/05/2017
County = 07 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *RFI CHECK*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = BURK 8164.13.00 06/09
05/17/2017 11:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:47
02 Std. Gas	0.082	11:48
03 Room Air	0.000	11:48
04 Std. Gas	0.082	11:49
05 Room Air	0.000	11:49
06 Std. Gas	0.082	11:49
07 Room Air	0.000	11:50

Lot No = 19615080A5
Cyl No = 18
Exp Date = 09/05/2017
County = 07 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = BURK 8164.13.00 06/09
05/17/2017 11:50

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:51
02 Std. Gas	0.081	11:51
03 Room Air	0.000	11:52
04 Std. Gas	0.082	11:52
05 Room Air	0.000	11:53
06 Std. Gas	0.082	11:53
07 Room Air	0.000	11:53

Lot No = 19615080A5
Cyl No = 18
Exp Date = 09/05/2017
County = 07 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = BURK 8164.13.00 06/09
05/17/2017 11:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:54
02 Std. Gas	0.083	11:54
03 Room Air	0.000	11:55
04 Std. Gas	0.083	11:55
05 Room Air	0.000	11:56
06 Std. Gas	0.082	11:56
07 Room Air	0.000	11:57

Lot No = 19615080A5
Cyl No = 18
Exp Date = 09/05/2017
County = 07 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080AC

Form 106-I8000