

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-005942

Location: POWERS LAKE P.D

- | | Check When Done |
|--|-----------------|
| A. Is the breath tube heated? | <u>✓</u> |
| B. Is the diagnostic check complete? | <u>✓</u> |
| C. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| D. Print test. (Attach test record.) | <u>✓</u> |
| E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR5713</u> Lot No. <u>201507F</u> Rep AC <u>0.020</u> | <u>✓</u> |
| F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3378</u> Lot No. <u>16060</u> Rep AC <u>0.300</u> | <u>✓</u> |
| G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN <u>DR3803</u> Lot No. <u>ICS4</u> Solution <u>0.10AC+0.05% ACETONE</u> | <u>✓</u> |
| H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) | <u>✓</u> |
| I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) <u>0.080</u> AC | |

Lot No. 1961508045 Cyl No. 17 Exp. Date 9/5/2017

Test 1 <u>0.078</u>	Test 4 <u>0.077</u>	Test 7 <u>0.078</u>
Test 2 <u>0.078</u>	Test 5 <u>0.077</u>	Test 8 <u>0.076</u>
Test 3 <u>0.078</u>	Test 6 <u>0.077</u>	Test 9 <u>0.078</u>

Average 0.077

J. Tank pressure, Level 3 Func D Sub G P: 1014 psi Reg 1000 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION

5/17/2017
Date
22 May 17
Date

Charles E. Ed
Field Inspector's Signature
Deb Sheaver
Reviewed by

Intoxilyzer Test Record and Checklist
 NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
 North Dakota Model 8000 SN 80-005942
 Location = PLPD 8164.13.00 06/09
 05/17/2017 09:49

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
 ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
 ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
 Start Test Sequence: DABACABA
 Display Prelim Rslt? Yes
 Display Third Digit? Yes
 Inhib Printer(Y/N)? No
 Display Volume? No
 Disable On Memfull? Yes
 # of Print Copies? 1
 Select Std (D/W/I)? Dry
 Standard Value? 0.080
 Standard Lot #? 19615080A5
 Standard Cyl #? 17
 Standard Expiration? 09/05/2017
 Oper No? 666666

Flow Cal. Date: 06/18/2013
 Slope 715
 Intercept -740740

IR Calibration Date: 05/20/2015
 3um 9um

	3um	9um
0th Coef(*100):	-32049	-32893
1st Coef(*100):	272377	139703
2nd Coef(*100):	2023	1232
H2O adj(mg/l*10k):	468	428

***** Printer Test End *****



Operator Signature
 CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005942
Location = PLPD 8164.13.00 06/09
05/17/2017 09:55

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:56
02 Std. Sol.	0.017	09:57
03 Room Air	0.000	09:57
04 Std. Sol.	0.017	09:58
05 Room Air	0.000	09:59
06 Std. Sol.	0.017	09:59
07 Room Air	0.000	10:00

08 Sim Temp = 34.0°C

Simul Ser No = DR5113
Std Sol No = 201507F
County = 07 Oper No. = 666666


Operator Signature
CHARLES EDER

Remarks: *Low AC*
 0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005942
Location = PLPD 8164.13.00 06/09
05/17/2017 10:02

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:03
02 Std. Sol.	0.306	10:04
03 Room Air	0.000	10:04
04 Std. Sol.	0.307	10:05
05 Room Air	0.000	10:05
06 Std. Sol.	0.308	10:06
07 Room Air	0.000	10:07

08 Sim Temp = 34.0°C

Simul Ser No = DR3378

Std Sol No = 16060

County = 07

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005942
Location = PLPD 8164.13.00 06/09
05/17/2017 10:09

Test	AC	Time
01 Room Air	0.000	10:10
02 *Subject Test	INT*	10:11
03 Room Air	0.000	10:11

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 19615080A5
Cyl No = 17
Expiration Date = 09/05/2017
County = 07 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent
0.10AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005942
Location = PLPD 8164.13.00 06/09
05/17/2017 10:12

Test	AC	Time
01 Diagnostic	OK	10:12
02 Room Air	0.000	10:13
03 *Subject Test	RFI*	10:13
04 Room Air	0.000	10:14

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 19615080A5
Cyl No = 17
Expiration Date = 09/05/2017
County = 07 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005942
Location = PLPD 8164.13.00 06/09
05/17/2017 09:50

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:51
02 Std. Gas	0.078	09:51
03 Room Air	0.000	09:51
04 Std. Gas	0.078	09:52
05 Room Air	0.000	09:52
06 Std. Gas	0.078	09:53
07 Room Air	0.000	09:53

Lot No = 19615080A5
Cyl No = 17
Exp Date = 09/05/2017
County = 07 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005942
Location = PLPD 8164.13.00 06/09
05/17/2017 10:14

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:15
02 Std. Gas	0.077	10:15
03 Room Air	0.000	10:15
04 Std. Gas	0.077	10:16
05 Room Air	0.000	10:16
06 Std. Gas	0.077	10:17
07 Room Air	0.000	10:17

Lot No = 19615080A5
Cyl No = 17
Exp Date = 09/05/2017
County = 07 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005942
Location = PLPD 8164.13.00 06/09
05/17/2017 10:17

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:18
02 Std. Gas	0.078	10:18
03 Room Air	0.000	10:19
04 Std. Gas	0.076	10:19
05 Room Air	0.000	10:20
06 Std. Gas	0.078	10:20
07 Room Air	0.000	10:21

Lot No = 19615080A5
Cyl No = 17
Exp Date = 09/05/2017
County = 07 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Calibration Check
0.080 AC

Form 106-I8000