

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-006490

Location: TIOGA P.D. (TIOG)

- | | Check When Done |
|---|-----------------|
| A. Is the breath tube heated? | <u>✓</u> |
| B. Is the diagnostic check complete? | <u>✓</u> |
| C. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| D. Print test. (Attach test record.) | <u>✓</u> |
| E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR5113</u> Lot No. <u>201507F</u> Rep AC <u>0.020</u> | <u>✓</u> |
| F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3378</u> Lot No. <u>16060</u> Rep AC <u>0.300</u> | <u>✓</u> |
| G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN <u>DR3803</u> Lot No. <u>ICS4</u> Solution <u>0.10 AC + 0.05%
ACETONE</u> | <u>✓</u> |
| H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) | <u>✓</u> |
| I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) <u>0.080</u> AC | |

Lot No. 19615080A5 Cyl No. 13 Exp. Date 9/5/2017

Test 1 <u>0.080</u>	Test 4 <u>0.081</u>	Test 7 <u>0.080</u>
Test 2 <u>0.080</u>	Test 5 <u>0.080</u>	Test 8 <u>0.081</u>
Test 3 <u>0.080</u>	Test 6 <u>0.080</u>	Test 9 <u>0.081</u>

Average 0.080

- J. Tank pressure, Level 3 Func D Sub G P: 997 psi Reg 1000 psi
- K. Remarks/Maintenance record: ANNUAL INSPECTION

5/17/2017
Date

Charles E. Eda
Field Inspector's Signature

22 May 17
Date

Deb Shanaver
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006490
Location = TIOG 8164.14.00 09/16
05/17/2017 08:14

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 19615080A5
Standard Cyl #? 13
Standard Expiration? 09/05/2017
Oper No? 666666

Flow Cal. Date: 08/18/2015
Slope 675
Intercept -635799

IR Calibration Date: 08/18/2015
 3um 9um

0th Coef(*100):	-18073	-15729
1st Coef(*100):	265337	133887
2nd Coef(*100):	3965	1560
H2O adj(mg/l*10k):	767	457

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006490
Location = TIOG 8164.14.00 09/16
05/17/2017 08:52

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:53
02 Std. Sol.	0.018	08:54
03 Room Air	0.000	08:54
04 Std. Sol.	0.019	08:55
05 Room Air	0.000	08:56
06 Std. Sol.	0.019	08:56
07 Room Air	0.000	08:57

08 Sim Temp = 34.0°C

Simul Ser No = DR5113

Std Sol No = 201507F

County = 53

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

Low AC

0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006490
Location = TIOG 8164.14.00 09/16
05/17/2017 08:46

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:46
02 Std. Sol.	0.307	08:47
03 Room Air	0.000	08:48
04 Std. Sol.	0.306	08:49
05 Room Air	0.000	08:49
06 Std. Sol.	0.307	08:50
07 Room Air	0.000	08:51


08 Sim Temp = 34.0°C

Simul Ser No = DR3378

Std Sol No = 16060

County = 53

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

HIGH AC

0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006490
Location = TIOG 8164.14.00 09/16
05/17/2017 08:16

Test	AC	Time
01 Diagnostic	OK	08:17
02 Room Air	0.000	08:17
03 *Subject Test	RFI*	08:18
04 Room Air	0.000	08:18

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 19615080A5
Cyl No = 13
Expiration Date = 09/05/2017
County = 53 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006490
Location = TIOG 8164.14.00 09/16
05/17/2017 08:23

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:23
02 Std. Gas	0.080	08:24
03 Room Air	0.000	08:24
04 Std. Gas	0.080	08:25
05 Room Air	0.000	08:25
06 Std. Gas	0.080	08:26
07 Room Air	0.000	08:26

Lot No = 19615080A5
Cyl No = 13
Exp Date = 09/05/2017
County = 53 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006490
Location = TIOG 8164.14.00 09/16
05/17/2017 08:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:27
02 Std. Gas	0.081	08:28
03 Room Air	0.000	08:28
04 Std. Gas	0.080	08:29
05 Room Air	0.000	08:29
06 Std. Gas	0.080	08:30
07 Room Air	0.000	08:30


Lot No = 19615080A5

Cyl No = 13

Exp Date = 09/05/2017

County = 53

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006490
Location = TIOG 8164.14.00 09/16
05/17/2017 08:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:59
02 Std. Gas	0.080	08:59
03 Room Air	0.000	09:00
04 Std. Gas	0.081	09:00
05 Room Air	0.000	09:00
06 Std. Gas	0.081	09:01
07 Room Air	0.000	09:01

Lot No = 19615080A5
Cyl No = 13
Exp Date = 09/05/2017
County = 53 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK

0.080 AC

Form 106-I8000