

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-004953

Location: LINT (Emmons County SO.)

- Check When Done
- A. Is the breath tube heated? ✓
  - B. Is the diagnostic check complete? ✓
  - C. Is the time, date, and year correct (re-set if necessary)? ✓
  - D. Print test. (Attach test record.) ✓
  - E. Low AC. Use  $\leq 0.03$  AC in ACA mode. (Attach test record.)  
Sim SN DR5113 Lot No 201507F Rep AC 0.020 ✓
  - F. High AC. Use  $\geq 0.25$  AC in ACA mode. (Attach test record.)  
Sim SN DR3378 Lot No. 16060 Rep AC 0.300 ✓
  - G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.  
(Attach test record.)  
Sim SN DR3803 Lot No. ICS 5 Solution 0.10AC + 0.05% ACETONE ✓
  - H. RFI check. Run CMS mode. Key radio during CMS test.  
(Attach test record.) ✓

I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.  
(Attach test record.) 0.080 AC

Lot No. 1961508045 Cyl No. 5 Exp. Date 9/5/2017

|                     |                     |                     |
|---------------------|---------------------|---------------------|
| Test 1 <u>0.081</u> | Test 4 <u>0.080</u> | Test 7 <u>0.080</u> |
| Test 2 <u>0.081</u> | Test 5 <u>0.080</u> | Test 8 <u>0.081</u> |
| Test 3 <u>0.081</u> | Test 6 <u>0.080</u> | Test 9 <u>0.081</u> |

Average 0.081

J. Tank pressure, Level 3 Func D Sub G P: 987 psi Reg 1000 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION

5/23/2017  
Date

Charles E. Lee  
Field Inspector's Signature

24 May 17  
Date

Roberta Muegu-Thomson  
Reviewed by

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer           Alcohol Analyzer  
North Dakota Model 8000           SN 80-004953  
Location = LINT                   8164.13.00 06/09  
05/23/2017                           13:21

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^\*()\_+?

Current Instrument Setup

Data Entry Mode:           Enabled  
Start Test Sequence:       DABACABA  
Display Prelim Rslt?       Yes  
Display Third Digit?       Yes  
Inhib Printer(Y/N)?       No  
Display Volume?           No  
Disable On Memfull?       Yes  
# of Print Copies?         1  
Select Std (D/W/I)?       Dry  
Standard Value?           0.080  
Standard Lot #?           19615080A5  
Standard Cyl #?           5  
Standard Expiration?       09/05/2017  
Oper No?                   666666

Flow Cal. Date:           01/19/2010  
Slope                      691  
Intercept                 -628068

IR Calibration Date:       04/07/2015  
                              3um                   9um

-----  
0th Coef(\*100):       -18160           -22754  
1st Coef(\*100):       254326           129076  
2nd Coef(\*100):       3670             1866  
H2O adj(mg/l\*10k):   414               530

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks:       **PRINT TEST**

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004953  
Location = LINT      8164.13.00 06/09  
05/23/2017      13:16

WET CAL CHECK

| Test         | AC    | Time  |
|--------------|-------|-------|
| 01 Room Air  | 0.000 | 13:16 |
| 02 Std. Sol. | 0.017 | 13:17 |
| 03 Room Air  | 0.000 | 13:18 |
| 04 Std. Sol. | 0.017 | 13:18 |
| 05 Room Air  | 0.000 | 13:19 |
| 06 Std. Sol. | 0.018 | 13:20 |
| 07 Room Air  | 0.000 | 13:20 |

08 Sim Temp = 34.0°C

Simul Ser No = DR5113  
Std Sol No = 201507F  
County = 15      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *Low AC*  
*0.020 AC*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004953  
Location = LINT      8164.13.00 06/09  
05/23/2017      13:10

WET CAL CHECK

| Test         | AC    | Time  |
|--------------|-------|-------|
| 01 Room Air  | 0.000 | 13:11 |
| 02 Std. Sol. | 0.298 | 13:11 |
| 03 Room Air  | 0.000 | 13:12 |
| 04 Std. Sol. | 0.300 | 13:13 |
| 05 Room Air  | 0.000 | 13:13 |
| 06 Std. Sol. | 0.302 | 13:14 |
| 07 Room Air  | 0.000 | 13:15 |


08 Sim Temp = 34.0°C

Simul Ser No = DR3378

Std Sol No = 16060

County = 15

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

HIGH AC

0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004953  
Location = LINT      8164.13.00 06/09  
05/23/2017      13:25

| Test             | AC    | Time  |
|------------------|-------|-------|
| 01 Room Air      | 0.000 | 13:26 |
| 02 *Subject Test | INT*  | 13:26 |
| 03 Room Air      | 0.000 | 13:27 |

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = INTERFERENT  
Dr. Lic. = ND/DIS821456  
Lot No = 19615080A5  
Cyl No = 5  
Expiration Date = 09/05/2017  
County = 15      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

*Interferent*  
*0.10AC + 0.05% ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004953  
Location = LINT                    8164.13.00 06/09  
05/23/2017                            12:57

| Test             | AC    | Time  |
|------------------|-------|-------|
| 01 Diagnostic    | OK    | 12:58 |
| 02 Room Air      | 0.000 | 12:58 |
| 03 *Subject Test | RFI*  | 12:59 |
| 04 Room Air      | 0.000 | 12:59 |

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI CHECK  
Dr. Lic. = ND/DIS821456  
Lot No = 19615080A5  
Cyl No = 5  
Expiration Date = 09/05/2017  
County = 15                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks: *RFI CHECK*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004953  
Location = LINT                    8164.13.00 06/09  
05/23/2017                            13:22

DRY CAL CHECK

| Test        | AC    | Time  |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 13:22 |
| 02 Std. Gas | 0.081 | 13:23 |
| 03 Room Air | 0.000 | 13:23 |
| 04 Std. Gas | 0.081 | 13:23 |
| 05 Room Air | 0.000 | 13:24 |
| 06 Std. Gas | 0.081 | 13:24 |
| 07 Room Air | 0.000 | 13:25 |

Lot No = 19615080A5  
Cyl No = 5  
Exp Date = 09/05/2017  
County = 15                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
                  0.080 AC


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004953  
Location = LINT      8164.13.00 06/09  
05/23/2017      13:27

DRY CAL CHECK

| Test        | AC    | Time  |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 13:28 |
| 02 Std. Gas | 0.080 | 13:28 |
| 03 Room Air | 0.000 | 13:28 |
| 04 Std. Gas | 0.080 | 13:29 |
| 05 Room Air | 0.000 | 13:29 |
| 06 Std. Gas | 0.080 | 13:30 |
| 07 Room Air | 0.000 | 13:30 |

Lot No = 19615080A5  
Cyl No = 5  
Exp Date = 09/05/2017  
County = 15      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004953  
Location = LINT      8164.13.00 06/09  
05/23/2017      13:31

DRY CAL CHECK

| Test        | AC    | Time  |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 13:31 |
| 02 Std. Gas | 0.080 | 13:32 |
| 03 Room Air | 0.000 | 13:32 |
| 04 Std. Gas | 0.081 | 13:32 |
| 05 Room Air | 0.000 | 13:33 |
| 06 Std. Gas | 0.081 | 13:33 |
| 07 Room Air | 0.000 | 13:34 |

Lot No = 19615080A5  
Cyl No = 5  
Exp Date = 09/05/2017  
County = 15      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080 AC