

CHAIN OF CUSTODY
BREATH ALCOHOL INSTRUMENT

Instrument: 8000 Serial Number: 80-006504

Agency being assigned instrument: Crime Lab

Date 25 May 17

Transferred from: Bollineau Co SO

Transferred to:

Name (print) Deb Shanaver

Name (print) Deb Shanaver

Signature Deb Shanaver

Signature Deb Shanaver

Agency Crime Lab

Agency Crime Lab

Reason for transfer/remarks:

Exchanged during annual inspection because it couldn't perform a wet bath solution test. Subject testing would not have been affected.

Or Shipped Via:

USPS _____ UPS _____ FedEx _____ Other _____

Date Sending Instrument: _____ / _____ / _____ Air Bill No. _____