

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-004946

Location: TOXL

Check When Done

- A. Is the breath tube heated? ✓
- B. Is the diagnostic check complete? ✓
- C. Is the time, date, and year correct (re-set if necessary)? ✓
- D. Print test. (Attach test record.) ✓
- E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN MP3003 Lot No. 16200 Rep AC 0.020 ✓
- F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN MP3070 Lot No. 16060 Rep AC 0.300 ✓
- G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN DR5135 Lot No. ICS 4 Solution 0.10 AC + 0.05% ACETONE ✓
- H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) ✓
- I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 17316080A2 Cyl No. 28 Exp. Date 8/5/2018

Test 1 0.081 Test 4 0.081 Test 7 0.080

Test 2 0.080 Test 5 0.081 Test 8 0.081

Test 3 0.081 Test 6 0.081 Test 9 0.081

Average 0.081

J. Tank pressure, Level 3 Func D Sub G P: 350 psi Reg 350 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION

5/30/2017
Date

Charles E. Egan
Field Inspector's Signature

30 May 17
Date

Deb Shanaver
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = TOXL 8164.14.00 09/16
05/26/2017 16:45

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 17316080A2
Standard Cyl #? 28
Standard Expiration? 08/05/2018
Oper No? 666666

Flow Cal. Date: 01/18/2010
Slope 717
Intercept -728279

IR Calibration Date: 04/23/2015

	3um	9um
0th Coef(*100):	-7851	-17700
1st Coef(*100):	259859	133464
2nd Coef(*100):	3769	1675
H2O adj(mg/l*10k):	294	408

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks: *PRINT TEST*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = TOXL 8164.14.00 09/16
05/26/2017 16:48

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:49
02 Std. Sol.	0.018	16:49
03 Room Air	0.000	16:50
04 Std. Sol.	0.019	16:51
05 Room Air	0.000	16:51
06 Std. Sol.	0.018	16:52
07 Room Air	0.000	16:52

08 Sim Temp = 34.0°C

Simul Ser No = MP3003

Std Sol No = 16200

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = TOXL 8164.14.00 09/16
05/30/2017 10:57

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:57
02 Std. Sol.	0.305	10:58
03 Room Air	0.000	10:58
04 Std. Sol.	0.305	10:59
05 Room Air	0.000	11:00
06 Std. Sol.	0.304	11:00
07 Room Air	0.000	11:01

08 Sim Temp = 34.0°C

Simul Ser No = MP3070
Std Sol No = 16060
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: HIGH AC
0.300 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = TOXL 8164.14.00 09/16
05/30/2017 11:05

Test	AC	Time
01 Room Air	0.000	11:06
02 *Subject Test	INT*	11:06
03 Room Air	0.000	11:07

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 17316080A2
Cyl No = 28
Expiration Date = 08/05/2018
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent
0.10AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = TOXL 8164.14.00 09/16
05/30/2017 11:08

Test	AC	Time
01 Diagnostic	OK	11:09
02 Room Air	0.000	11:10
03 *Subject Test	RFI*	11:10
04 Room Air	0.000	11:10

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 17316080A2
Cyl No = 28
Expiration Date = 08/05/2018
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *RFI CHECK*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = TOXL 8164.14.00 09/16
05/30/2017 11:11

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:11
02 Std. Gas	0.081	11:12
03 Room Air	0.000	11:12
04 Std. Gas	0.080	11:12
05 Room Air	0.000	11:13
06 Std. Gas	0.081	11:13
07 Room Air	0.000	11:14

Lot No = 17316080A2
Cyl No = 28
Exp Date = 08/05/2018
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK

0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = TOXL 8164.14.00 09/16
05/30/2017 11:14

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:15
02 Std. Gas	0.081	11:15
03 Room Air	0.000	11:16
04 Std. Gas	0.081	11:16
05 Room Air	0.000	11:16
06 Std. Gas	0.081	11:17
07 Room Air	0.000	11:17

Lot No = 17316080A2
Cyl No = 28
Exp Date = 08/05/2018
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = TOXL 8164.14.00 09/16
05/30/2017 11:18

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:18
02 Std. Gas	0.080	11:18
03 Room Air	0.000	11:19
04 Std. Gas	0.081	11:19
05 Room Air	0.000	11:20
06 Std. Gas	0.081	11:20
07 Room Air	0.000	11:21

Lot No = 17316080A2
Cyl No = 28
Exp Date = 08/05/2018
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC