

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-006496

Location: BWLM  
Burman CO SO

- |   | Check When Done |
|---|-----------------|
| A. Is the breath tube heated?   | <u>X</u>        |
| B. Is the diagnostic check complete?  | <u>X</u>        |
| C. Is the time, date, and year correct (re-set if necessary)?   | <u>X</u>        |
| D. Print test. (Attach test record.)  | <u>X</u>        |
| E. Low AC. Use $\leq 0.03$ AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5113</u> Lot No. <u>201507F</u> Rep AC <u>0.020AC</u>   | <u>X</u>        |
| F. High AC. Use $\geq 0.25$ AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR3378</u> Lot No. <u>16060</u> Rep AC <u>0.300AC</u>  | <u>X</u>        |
| G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.<br>(Attach test record.)<br>Sim SN <u>DR3803</u> Lot No. <u>1CS5</u> Solution <u>0.100AC + 0.05%<br/>Acetone</u> | <u>X</u>        |
| H. RFI check. Run CMS mode. Key radio during CMS test.<br>(Attach test record.)   | <u>X</u>        |
| I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.<br>(Attach test record.) <u>0.080</u> AC   |                 |

Lot No. <u>01316080A2</u>	Cyl No. <u>25</u>	Exp. Date <u>5 Feb 18</u>
Test 1 <u>0.080</u>	Test 4 <u>0.081</u>	Test 7 <u>0.080</u>
Test 2 <u>0.080</u>	Test 5 <u>0.081</u>	Test 8 <u>0.080</u>
Test 3 <u>0.080</u>	Test 6 <u>0.080</u>	Test 9 <u>0.081</u>
Average <u>0.080</u>		

J. Tank pressure, Level 3 Func D Sub G P: 976 psi Reg 975 psi

K. Remarks/Maintenance record: Annual inspection, replaced O-ring on breath hose.

30 May 17  
Date

31 May 17  
Date

Roberta Hume - Humes  
Field Inspector's Signature

Deb Shanaver  
Reviewed by

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006496  
Location = BOWM                    8164.13.00 06/09  
05/30/2017                            11:44

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            01316080A2  
Standard Cyl #?            25  
Standard Expiration?       02/05/2018  
Oper No?                    888888

Flow Cal. Date:            08/18/2015  
Slope                      673  
Intercept                  -654167

IR Calibration Date:       08/18/2015  
                                  3um                    9um

-----  
0th Coef(\*100):        -22081                -15963  
1st Coef(\*100):        278490                136818  
2nd Coef(\*100):        2086                   1282  
H2O adj(mg/l\*10k):    598                    386

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: *Print Test*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006496  
Location = BOWM      8164.13.00 06/09  
05/30/2017      12:07

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:07
02 Std. Sol.	0.019	12:09
03 Room Air	0.000	12:09
04 Std. Sol.	0.020	12:10
05 Room Air	0.000	12:11
06 Std. Sol.	0.020	12:11
07 Room Air	0.000	12:12

08 Sim Temp = 34.0°C

Simul Ser No = DR5113  
Std Sol No = 201507F  
County = 06      Oper No. = 888888

  
Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: LOW AC  
0.020 AC

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006496  
Location = BOWM      8164.13.00 06/09  
05/30/2017      12:15

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:15
02 Std. Sol.	0.298	12:16
03 Room Air	0.000	12:17
04 Std. Sol.	0.298	12:18
05 Room Air	0.000	12:18
06 Std. Sol.	0.300	12:19
07 Room Air	0.000	12:20


08 Sim Temp = 34.0°C

Simul Ser No = DR3378

Std Sol No = 16060

County = 06

Oper No. = 888888



Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: High AC  
0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006496  
Location = BOWM                    8164.13.00 06/09  
05/30/2017                            11:45

Test	AC	Time
01 Room Air	0.000	11:45
02 *Subject Test	INT*	11:46
03 Room Air	0.000	11:47

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = INTERFERENT  
Dr. Lic. = ND/DIS821456  
Lot No = 01316080A2  
Cyl No = 25  
Expiration Date = 02/05/2018  
County = 06                            Oper No. = 888888

  
Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: Interferent Check  
0.100 AC + 0.05% Acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006496  
Location = BOWM                    8164.13.00 06/09  
05/30/2017                            13:47

Test	AC	Time
01 Diagnostic	OK	13:48
02 Room Air	0.000	13:49
03 *Subject Test	RFI*	13:49
04 Room Air	0.000	13:50

\*Invalid Test  
  Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI CHECK  
Dr. Lic. = ND/DIS821456  
Lot No = 01316080A2  
Cyl No = 25  
Expiration Date = 02/05/2018  
County = 06                            Oper No. = 888888

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

  
Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: *RFI Check*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006496  
Location = BOWM      8164.13.00 06/09  
05/30/2017      11:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:48
02 Std. Gas	0.080	11:48
03 Room Air	0.000	11:49
04 Std. Gas	0.080	11:49
05 Room Air	0.000	11:50
06 Std. Gas	0.080	11:50
07 Room Air	0.000	11:51

Lot No = 01316080A2  
Cyl No = 25  
Exp Date = 02/05/2018  
County = 06      Oper No. = 888888

  
Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: Calibration Check  
0.080AC

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006496  
Location = BOWM      8164.13.00 06/09  
05/30/2017      11:51

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:52
02 Std. Gas	0.081	11:52
03 Room Air	0.000	11:53
04 Std. Gas	0.081	11:53
05 Room Air	0.000	11:54
06 Std. Gas	0.080	11:54
07 Room Air	0.000	11:55

Lot No = 01316080A2  
Cyl No = 25  
Exp Date = 02/05/2018  
County = 06      Oper No. = 888888

  
Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: Calibration Check  
0.080 AC




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006496  
Location = BOWM      8164.13.00 06/09  
05/30/2017      11:55

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:56
02 Std. Gas	0.080	11:56
03 Room Air	0.000	11:57
04 Std. Gas	0.080	11:57
05 Room Air	0.000	11:58
06 Std. Gas	0.081	11:58
07 Room Air	0.000	11:59

Lot No = 01316080A2  
Cyl No = 25  
Exp Date = 02/05/2018  
County = 06      Oper No. = 888888

  
Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: Calibration Check  
0.080 AC

Form 106-I8000