

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-006498

Location: Minot PD

- | | Check When Done |
|---|-----------------|
| A. Is the breath tube heated? | <u>X</u> |
| B. Is the diagnostic check complete? | <u>X</u> |
| C. Is the time, date, and year correct (re-set if necessary)? | <u>X</u> |
| D. Print test. (Attach test record.) | <u>X</u> |
| E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR5113</u> Lot No. <u>201507F</u> Rep AC <u>0.020AC</u> | <u>X</u> |
| F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3378</u> Lot No. <u>16060</u> Rep AC <u>0.300AC</u> | <u>X</u> |
| G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN <u>DR3803</u> Lot No. <u>1CS5</u> Solution <u>0.100AC + 0.05% Acetone</u> | <u>X</u> |
| H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) | <u>X</u> |

- I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 09216D80A3 Cyl No. 28 Exp. Date 5 May 18
~~5 Mar 18~~ Ren

Test 1	<u>0.080</u>	Test 4	<u>0.080</u>	Test 7	<u>0.081</u>
Test 2	<u>0.080</u>	Test 5	<u>0.081</u>	Test 8	<u>0.081</u>
Test 3	<u>0.081</u>	Test 6	<u>0.080</u>	Test 9	<u>0.080</u>

Average 0.080

- J. Tank pressure, Level 3 Func D Sub G P: 177 psi Reg ~200 psi

- K. Remarks/Maintenance record: Annual inspection. Re-calibrated
tan-BW tank sensor.

16 June 17
Date

06 June 17
Date

Roberta Muga-Timmo
Field Inspector's Signature

D. Shanaver
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006498
Location = MIPD 8164.14.00 09/16
06/06/2017 12:35

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 09216080A3
Standard Cyl #? 28
Standard Expiration? 05/05/2018
Oper No? 888888

Flow Cal. Date: 08/18/2015
Slope 669
Intercept -637469

IR Calibration Date: 08/18/2015
3um 9um

	3um	9um
0th Coef(*100):	-28161	-22508
1st Coef(*100):	286657	143843
2nd Coef(*100):	2341	1281
H2O adj(mg/l*10k):	776	451

***** Printer Test End *****



Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: *Print Test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006498
Location = MIPD 8164.14.00 09/16
06/06/2017 12:38

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:39
02 Std. Sol.	0.020	12:40
03 Room Air	0.000	12:40
04 Std. Sol.	0.019	12:41
05 Room Air	0.000	12:41
06 Std. Sol.	0.020	12:42
07 Room Air	0.000	12:43

08 Sim Temp = 34.0°C

Simul Ser No = DR5113
Std Sol No = 201507F
County = 51 Oper No. = 888888



Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks:

LOW AC
0.020AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006498
Location = MIPD 8164.14.00 09/16
06/06/2017 12:45

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:45
02 Std. Sol.	0.304	12:46
03 Room Air	0.000	12:46
04 Std. Sol.	0.305	12:47
05 Room Air	0.000	12:48
06 Std. Sol.	0.305	12:48
07 Room Air	0.000	12:49


08 Sim Temp = 34.0°C

Simul Ser No = DR3378

Std Sol No = 16060

County = 51

Oper No. = 888888



Operator Signature

ROBERTA GRIEGER-NIMMO

Remarks:

High AC
0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006498
Location = MIPD 8164.14.00 09/16
06/06/2017 12:54

Test	AC	Time
01 Room Air	0.000	12:54
02 *Subject Test	INT*	12:55
03 Room Air	0.000	12:56

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 09216080A3
Cyl No = 28
Expiration Date = 05/05/2018
County = 51 Oper No. = 888888



Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: *Interferent Check*
0.100AC + 0.05% Acetone

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006498
Location = MIPD 8164.14.00 09/16
06/06/2017 12:57

Test	AC	Time
01 Diagnostic	OK	12:58
02 Room Air	0.000	12:58
03 *Subject Test	RFI*	12:58
04 Room Air	0.000	12:59

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI
Dr. Lic. = ND/DIS821456
Lot No = 09216080A3
Cyl No = 28
Expiration Date = 05/05/2018
County = 51 Oper No. = 888888

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: *RFI Check*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006498
Location = MIPD 8164.14.00 09/16
06/06/2017 12:59

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:00
02 Std. Gas	0.080	13:00
03 Room Air	0.000	13:00
04 Std. Gas	0.080	13:01
05 Room Air	0.000	13:01
06 Std. Gas	0.081	13:02
07 Room Air	0.000	13:02

Lot No = 09216080A3
Cyl No = 28
Exp Date = 05/05/2018
County = 51 Oper No. = 888888


Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: Calibration Check
0.080AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006498
Location = MIPD 8164.14.00 09/16
06/06/2017 13:02

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:03
02 Std. Gas	0.080	13:03
03 Room Air	0.000	13:04
04 Std. Gas	0.081	13:04
05 Room Air	0.000	13:05
06 Std. Gas	0.080	13:05
07 Room Air	0.000	13:06

Lot No = 09216080A3
Cyl No = 28
Exp Date = 05/05/2018
County = 51 Oper No. = 888888


Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: Calibration Check
0.080 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006498
Location = MIPD 8164.14.00 09/16
06/06/2017 13:06

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:06
02 Std. Gas	0.081	13:07
03 Room Air	0.000	13:07
04 Std. Gas	0.081	13:08
05 Room Air	0.000	13:08
06 Std. Gas	0.080	13:09
07 Room Air	0.000	13:09

Lot No = 09216080A3
Cyl No = 28
Exp Date = 05/05/2018
County = 51 Oper No. = 888888



Operator Signature
ROBERTA GRIEGER-NIMMO

Remarks: Calibration Check
0.080 AC

Form 106-I8000