

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-006503

Location: ToxL

- Check When Done
- A. Is the breath tube heated? ✓
- B. Is the diagnostic check complete? ✓
- C. Is the time, date, and year correct (re-set if necessary)? ✓
- D. Print test. (Attach test record.) ✓
- E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN MP3003 Lot No. 16200 Rep AC 0.020 ✓
- F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN MP3070 Lot No. 16060 Rep AC 0.300 ✓
- G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN MP3071 Lot No. ICS5 Solution 0.10AC + 0.05% ACETONE ✓
- H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) ✓
- I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 03417080A3 Cyl No. 27 Exp. Date 3/5/2019

Test 1 0.081 Test 4 0.081 Test 7 0.081

Test 2 0.081 Test 5 0.082 Test 8 0.081

Test 3 0.081 Test 6 0.081 Test 9 0.081

Average 0.081

J. Tank pressure, Level 3 Func D Sub G P: 996 psi Reg 1000 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION

6/14/2017
Date

6/15/2017
Date

Charles E. Eda
Field Inspector's Signature

Roberta Huger-Timm
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006503
Location = TOXL 8164.14.00 09/16
06/14/2017 10:48

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 03417080A3
Standard Cyl #? 27
Standard Expiration? 03/05/2019
Oper No? 666666

Flow Cal. Date: 08/20/2015
Slope 663
Intercept -633300

IR Calibration Date: 06/12/2017

	3um	9um
0th Coef(*100):	-6741	-15632
1st Coef(*100):	256583	132194
2nd Coef(*100):	3028	1131
H2O adj(mg/l*10k):	354	358

***** Printer Test End *****


Operator Signature
CHARLES EDER

Remarks: PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006503
Location = TOXL 8164.14.00 09/16
06/14/2017 10:21

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:22
02 Std. Sol.	0.019	10:23
03 Room Air	0.000	10:23
04 Std. Sol.	0.019	10:24
05 Room Air	0.000	10:25
06 Std. Sol.	0.019	10:25
07 Room Air	0.000	10:26

08 Sim Temp = 34.0°C

Simul Ser No = MP3003

Std Sol No = 16200

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006503
Location = TOXL 8164.14.00 09/16
06/14/2017 10:33

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:33
02 Std. Sol.	0.304	10:34
03 Room Air	0.000	10:35
04 Std. Sol.	0.303	10:35
05 Room Air	0.000	10:36
06 Std. Sol.	0.303	10:37
07 Room Air	0.000	10:37

08 Sim Temp = 34.0°C

Simul Ser No = MP3070
Std Sol No = 16060
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: HIGH AC
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006503
Location = TOXL 8164.14.00 09/16
06/14/2017 10:51

Test	AC	Time
01 Room Air	0.000	10:51
02 *Subject Test	INT*	10:52
03 Room Air	0.000	10:52

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 03417080A3
Cyl No = 27
Expiration Date = 03/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Interferent Check*
0.10 AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006503
Location = TOXL 8164.14.00 09/16
06/14/2017 10:54

Test	AC	Time
01 Diagnostic	OK	10:54
02 Room Air	0.000	10:55
03 *Subject Test	RFI*	10:55
04 Room Air	0.000	10:56

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 03417080A3
Cyl No = 27
Expiration Date = 03/05/2019
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006503
Location = TOXL 8164.14.00 09/16
06/14/2017 10:57

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:57
02 Std. Gas	0.081	10:57
03 Room Air	0.000	10:58
04 Std. Gas	0.081	10:58
05 Room Air	0.000	10:59
06 Std. Gas	0.081	10:59
07 Room Air	0.000	11:00

Lot No = 03417080A3
Cyl No = 27
Exp Date = 03/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *CALIBRATION CHECK*
0.080 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006503
Location = TOXL 8164.14.00 09/16
06/14/2017 11:00

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:01
02 Std. Gas	0.081	11:01
03 Room Air	0.000	11:02
04 Std. Gas	0.082	11:02
05 Room Air	0.000	11:03
06 Std. Gas	0.081	11:03
07 Room Air	0.000	11:04

Lot No = 03417080A3
Cyl No = 27
Exp Date = 03/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006503
Location = TOXL 8164.14.00 09/16
06/14/2017 11:04

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:05
02 Std. Gas	0.081	11:05
03 Room Air	0.000	11:05
04 Std. Gas	0.081	11:06
05 Room Air	0.000	11:06
06 Std. Gas	0.081	11:07
07 Room Air	0.000	11:07

Lot No = 03417080A3
Cyl No = 27
Exp Date = 03/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC