

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-004940

Location: FTOT (SPIRIT LAKE AGENCY, FORT TOTTEN)

- Check When Done
- A. Is the breath tube heated? ✓
 - B. Is the diagnostic check complete? ✓
 - C. Is the time, date, and year correct (re-set if necessary)? ✓
 - D. Print test. (Attach test record.) ✓
 - E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5113 Lot No. 201507F Rep AC 0.020 ✓
 - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3378 Lot No. 16060 Rep AC 0.300 ✓
 - G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN DR3803 Lot No. ICS5 Solution 0.10 AC + 0.05% ACETONE ✓
 - H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) ✓
 - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 09216080A3 Cyl No. 26 Exp. Date 5/5/2018

Test 1 <u>0.081</u>	Test 4 <u>0.081</u>	Test 7 <u>0.081</u>
Test 2 <u>0.081</u>	Test 5 <u>0.081</u>	Test 8 <u>0.080</u>
Test 3 <u>0.082</u>	Test 6 <u>0.082</u>	Test 9 <u>0.081</u>

Average 0.081

J. Tank pressure, Level 3 Func D Sub G P: 971 psi Reg 975 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION

6/15/2017
Date

16 June 2017
Date

Charles E. Ehl
Field Inspector's Signature

Deb Shanaver
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = FTOT 8164.14.00 09/16
06/15/2017 10:44

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 09216080A3
Standard Cyl #? 26
Standard Expiration? 05/05/2018
Oper No? 666666

Flow Cal. Date: 01/18/2010
Slope 720
Intercept -746294

IR Calibration Date: 06/01/2015
 3um 9um

0th Coef(*100): -14429 -15122
1st Coef(*100): 263440 134864
2nd Coef(*100): 2296 1357
H2O adj(mg/l*10k): 371 332

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = FTOT 8164.14.00 09/16
06/15/2017 11:11

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:11
02 Std. Sol.	0.020	11:12
03 Room Air	0.000	11:13
04 Std. Sol.	0.020	11:13
05 Room Air	0.000	11:14
06 Std. Sol.	0.019	11:15
07 Room Air	0.000	11:15

08 Sim Temp = 34.0°C

Simul Ser No = DR5113
Std Sol No = 201507F
County = 62 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Low AC*
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = FTOT 8164.14.00 09/16
06/15/2017 11:05

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:05
02 Std. Sol.	0.296	11:06
03 Room Air	0.000	11:07
04 Std. Sol.	0.298	11:07
05 Room Air	0.000	11:08
06 Std. Sol.	0.298	11:09
07 Room Air	0.000	11:09

08 Sim Temp = 34.0°C

Simul Ser No = DR3378

Std Sol No = 16060

County = 62

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = FTOT 8164.14.00 09/16
06/15/2017 10:58

Test	AC	Time
01 Room Air	0.000	10:59
02 *Subject Test	INT*	11:00
03 Room Air	0.000	11:01

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, SPIRIT THE
Sub DOB = 10/31/2005
Sub Sex = Male Weight = 280
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS05123456
Lot No = 09216080A3
Cyl No = 26
Expiration Date = 05/05/2018
County = 62 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10 AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = FTOT 8164.14.00 09/16
06/15/2017 11:01

Test	AC	Time
01 Diagnostic	OK	11:02
02 Room Air	0.000	11:03
03 *Subject Test	RFI*	11:03
04 Room Air	0.000	11:03

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, SPIRIT THE
Sub DOB = 10/31/2005
Sub Sex = Male Weight = 280
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS05123456
Lot No = 09216080A3
Cyl No = 26
Expiration Date = 05/05/2018
County = 62 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = FTOT 8164.14.00 09/16
06/15/2017 10:51

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:52
02 Std. Gas	0.081	10:52
03 Room Air	0.000	10:53
04 Std. Gas	0.081	10:53
05 Room Air	0.000	10:53
06 Std. Gas	0.082	10:54
07 Room Air	0.000	10:54

Lot No = 09216080A3
Cyl No = 26
Exp Date = 05/05/2018
County = 62 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = FTOT 8164.14.00 09/16
06/15/2017 10:55

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:55
02 Std. Gas	0.081	10:56
03 Room Air	0.000	10:56
04 Std. Gas	0.081	10:56
05 Room Air	0.000	10:57
06 Std. Gas	0.082	10:57
07 Room Air	0.000	10:58

Lot No = 09216080A3
Cyl No = 26
Exp Date = 05/05/2018
County = 62 Oper No. = 666666


Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = FTOT 8164.14.00 09/16
06/15/2017 11:16

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:16
02 Std. Gas	0.081	11:17
03 Room Air	0.000	11:17
04 Std. Gas	0.080	11:17
05 Room Air	0.000	11:18
06 Std. Gas	0.081	11:18
07 Room Air	0.000	11:19

Lot No = 09216080A3
Cyl No = 26
Exp Date = 05/05/2018
County = 62 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000