



**ETHANOL GAS STANDARD CYLINDER REPORT**

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>Melby, Robert M.</i>		
Location <i>"CROS"</i>	Intoxilyzer® Serial Number <i>80-005957</i>	
Gas Lot Number <i>17316080A2</i>	Gas Cylinder Number <i>17</i>	Gas Expiration Date <i>08-05-2018</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).

Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).

- A. *0.079* AC
- B. *0.078* AC
- C. *0.078* AC

- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature <i>Robert M. Melby</i>	Date <i>09-01-2017</i>
Reviewed By (Crime Laboratory Use Only) <i>Roberta Sanchez-Nimmo</i>	Date <i>09-07-2017</i>


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005957  
Location = CROS      8164.14.00 09/16  
09/01/2017      15:17

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:17
02 Std. Gas	0.079	15:17
03 Room Air	0.000	15:18
04 Std. Gas	0.078	15:18
05 Room Air	0.000	15:19
06 Std. Gas	0.078	15:19
07 Room Air	0.000	15:20

Lot No = 17316080A2  
Cyl No = 17  
Exp Date = 08/05/2018  
County = 12      Oper No. = 100011


  
\_\_\_\_\_  
Operator Signature  
ROBERT MELBY

Remarks:

Test good

Form 106-I8000

The undersigned, having custody of the original record, certifies that the information hereon is a true and correct copy of the original document maintained as part of the activities of this office.

  
\_\_\_\_\_  
Signature

09-01-2017  
Date