



**ETHANOL GAS STANDARD CYLINDER REPORT**  
 OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION  
 Toxicology Section/Breath Alcohol Program  
 SFN 59282 (10-2013)

Chemical Test Operator Name (Print) HUNTER RHOADES		
Location 295 OWENS ST, MANNING, ND, 58042 (DCSA) MANNING-SHERIFF'S OFFICE	Intoxilyzer® Serial Number 89-005955	
Gas Lot Number 03417080A3	Gas Cylinder Number CYL 031 / 03417080A3 (LOT)	Gas Expiration Date 03/05/2019

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).  
 A. 0.083 AC  
 B. 0.082 AC  
 C. 0.082 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:

- Form 105-G A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
- and ACA B. ACA Test Record.
- received 7 Sep 17
- RSN

Chemical Test Operator Signature <i>Hunter Rhoades</i>	Date 9/7/17
Reviewed By (Crime Laboratory Use Only) <i>Roberta Suegi-Tummo</i>	Date 7 Sep 17


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005955  
Location = DUNN      8164.14.00 09/16  
09/07/2017      05:13

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	05:14
02 Std. Gas	0.083	05:14
03 Room Air	0.000	05:15
04 Std. Gas	0.082	05:15
05 Room Air	0.000	05:16
06 Std. Gas	0.082	05:16
07 Room Air	0.000	05:16

Lot No = 03417080A3  
Cyl No = 31  
Exp Date = 03/05/2019  
County = 13      Oper No. = 130635


  
Operator Signature  
HUNTER RHOADES

Remarks: COMPLETED W/ Form 105-G ; NEW GAS CYLINDER INSTALL.

Form 106-I8000

I certify that the attached is a true and correct copy of the intoxilyzer test record.

Print Name/Badge #: HUNTER RHOADES/3261

Signature:  3261

Date: 9/7/17 Dept: DCSO