



ETHANOL GAS STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (10-2013)

Chemical Test Operator Name (Print) GALLAGHER, JONATHAN		
Location HARV	Intoxilyzer® Serial Number 80-006500	
Gas Lot Number 03417080A3	Gas Cylinder Number 1	Gas Expiration Date 03/05/2019

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. 0. 0 8 0 AC
 - B. 0. 0 8 1 AC
 - C. 0. 0 8 1 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:

ACA Test record
and Form 105-G
SFN59282 rec'd
at Crime Lab
10/2/17 - RJN

- A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
- B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i>	Date 09/16/2017
Reviewed By (Crime Laboratory Use Only) <i>[Signature]</i>	Date 10/02/2017


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006500
Location = HARV 8164.13.00 06/09
09/16/2017 17:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:08
02 Std. Gas	0.080	17:08
03 Room Air	0.000	17:09
04 Std. Gas	0.081	17:09
05 Room Air	0.000	17:10
06 Std. Gas	0.081	17:10
07 Room Air	0.000	17:11

Lot No = 03417080A3
Cyl No = 1
Exp Date = 03/05/2019
County = 52 Oper No. = 131749




Operator Signature
JONATHAN GALLAGHER

Remarks:

Form 106-I8000

The undersigned, having custody of the original public record, certifies that the information herein is a true and correct copy of the original document maintained as a part of the activities of this department for


Signature
Deputy Sheriff
Title
WCSO
Office/Agency