



ETHANOL GAS STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>BENNETT BITE</i>		
Location <i>Stutsman LEC Jail</i>	Intoxilyzer® Serial Number <i>80-006499</i>	
Gas Lot Number <i>03417080A3</i>	Gas Cylinder Number <i>4</i>	Gas Expiration Date <i>3/05/2019</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. *0.081* AC
 B. *0.081* AC
 C. *0.081* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Bennett Bite</i>	Date <i>9/21/17</i>
Reviewed By (Crime Laboratory Use Only) <i>Robertallegre Timmer</i>	Date <i>9/21/17</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006499
Location = JAME 8164.14.00 09/16
09/21/2017 09:02

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:03
02 Std. Gas	0.081	09:03
03 Room Air	0.000	09:03
04 Std. Gas	0.081	09:04
05 Room Air	0.000	09:04
06 Std. Gas	0.081	09:05
07 Room Air	0.000	09:05

Lot No = 03417080A3
Cyl No = 4
Exp Date = 03/05/2019
County = 47 Oper No. = 130997



Operator Signature
BENNETT BITZ

Remarks: