



ETHANOL GAS STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>James Christoff</i>		
Location <i>oliver County Sheriff's Office</i>	Intoxilyzer® Serial Number <i>80-005956</i>	
Gas Lot Number <i>20016080A2</i>	Gas Cylinder Number <i>CYL 039</i>	Gas Expiration Date <i>9/5/2018</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. *0.080* AC
 B. *0.080* AC
 C. *0.080* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:

Form 105-G and ACA Test Record were received at Crime Lab 10/11/17 RSN.

- A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
- B. ACA Test Record.

Chemical Test Operator Signature <i>James Christoff</i>	Date <i>10/10/17</i>
Reviewed By (Crime Laboratory Use Only) <i>Roberta Huegel-Turner</i>	Date <i>10/11/17</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = OLIV 8164.14.00 09/16
10/10/2017 17:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:38
02 Std. Gas	0.080	17:38
03 Room Air	0.000	17:39
04 Std. Gas	0.080	17:39
05 Room Air	0.000	17:40
06 Std. Gas	0.080	17:40
07 Room Air	0.000	17:40

Lot No = 20016080A2
Cyl No = 039
Exp Date = 09/05/2018
County = 33 Oper No. = 131189



Operator Signature
JAMES CHRISTOFF

Remarks: