



ETHANOL GAS STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>Duisey SHAWON</i> <i>020632</i>		
Location <i>TIOGA PD.</i>		Intoxilyzer® Serial Number <i>80-006490</i>
Gas Lot Number <i>03417080A3</i>	Gas Cylinder Number <i>8</i>	Gas Expiration Date <i>03/05/2019</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. *0.083* AC
 B. *0.082* AC
 C. *0.082* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i>	Date <i>10/31/17</i>
Reviewed By (Crime Laboratory Use Only) <i>Roberta Hughes-Nimmo</i>	Date <i>11/1/17</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006490
Location = TIOG 8164.14.00 09/16
10/31/2017 18:46

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:47
02 Std. Gas	0.083	18:47
03 Room Air	0.000	18:48
04 Std. Gas	0.082	18:48
05 Room Air	0.000	18:49
06 Std. Gas	0.082	18:49
07 Room Air	0.000	18:50

Lot No = 03417080A3
Cyl No = 8
Exp Date = 03/05/2019
County = 53 Oper No. = 020632



Operator Signature
SHANNON DUISEN

Remarks: