

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-005954

Location: DEVL

(LAKE REGION
LAW ENFORCEMENT
CENTER)

Check When Done

- A. Is the breath tube heated? ✓
- B. Is the diagnostic check complete? ✓
- C. Is the time, date, and year correct (re-set if necessary)? ✓
- D. Print test. (Attach test record.) ✓
- E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5113 Lot No. 201507F Rep AC 0.020 ✓
- F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3378 Lot No. 16060 Rep AC 0.300 ✓
- G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN DR3803 Lot No. ICS5 Solution 0.10 AC + 0.05% ACETONE ✓
- H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) ✓
- I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 09216080A3 Cyl No. 25 Exp. Date 5/5/2018
Test 1 0.080 Test 4 0.079 Test 7 0.079
Test 2 0.079 Test 5 0.080 Test 8 0.079
Test 3 0.079 Test 6 0.079 Test 9 0.079
Average 0.079

J. Tank pressure, Level 3 Func D Sub G P: 696 psi Reg 700 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION

6-15-2017

Date

16 June 2017 ✓

Date

Charles E. Ed
Field Inspector's Signature

Deb Shanover
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = DEVL 8164.14.00 09/16
06/15/2017 11:51

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 09216080A3
Standard Cyl #? 25
Standard Expiration? 05/05/2018
Oper No? 666666

Flow Cal. Date: 05/09/2014
Slope 647
Intercept -668435

IR Calibration Date: 05/12/2015
 3um 9um

0th Coef(*100): -16063 -18509
1st Coef(*100): 275605 134398
2nd Coef(*100): 3207 1457
H2O adj(mg/l*10k): 331 423

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks: PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = DEVL 8164.14.00 09/16
06/15/2017 11:53

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:53
02 Std. Sol.	0.017	11:54
03 Room Air	0.000	11:55
04 Std. Sol.	0.018	11:55
05 Room Air	0.000	11:56
06 Std. Sol.	0.017	11:57
07 Room Air	0.000	11:57

08 Sim Temp = 34.0°C

Simul Ser No = DR5113

Std Sol No = 201507F

County = 36

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC

0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = DEVL 8164.14.00 09/16
06/15/2017 12:20

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:21
02 Std. Sol.	0.290	12:21
03 Room Air	0.000	12:22
04 Std. Sol.	0.290	12:23
05 Room Air	0.000	12:23
06 Std. Sol.	0.292	12:24
07 Room Air	0.000	12:25

08 Sim Temp = 34.0°C

Simul Ser No = DR3378

Std Sol No = 16060

County = 36

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = DEVL 8164.14.00 09/16
06/15/2017 12:05

Test	AC	Time
01 Room Air	0.000	12:06
02 *Subject Test	INT*	12:07
03 Room Air	0.000	12:07

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, SPIRIT THE
Sub DOB = 10/31/2005
Sub Sex = Male Weight = 280
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS05123456
Lot No = 09216080A3
Cyl No = 25
Expiration Date = 05/05/2018
County = 36 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Interferent Check*
0.10 AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = DEVL 8164.14.00 09/16
06/15/2017 12:08

Test	AC	Time
01 Diagnostic	OK	12:10
02 Room Air	0.000	12:10
03 *Subject Test	RFI*	12:10
04 Room Air	0.000	12:11

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, SPIRIT THE
Sub DOB = 10/31/2005
Sub Sex = Male Weight = 280
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS05123456
Lot No = 09216080A3
Cyl No = 25
Expiration Date = 05/05/2018
County = 36 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = DEVL 8164.14.00 09/16
06/15/2017 12:11

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:12
02 Std. Gas	0.080	12:12
03 Room Air	0.000	12:13
04 Std. Gas	0.079	12:13
05 Room Air	0.000	12:14
06 Std. Gas	0.079	12:14
07 Room Air	0.000	12:15

Lot No = 09216080A3
Cyl No = 25
Exp Date = 05/05/2018
County = 36 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK

0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = DEVL 8164.14.00 09/16
06/15/2017 12:15

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:15
02 Std. Gas	0.079	12:16
03 Room Air	0.000	12:16
04 Std. Gas	0.080	12:17
05 Room Air	0.000	12:17
06 Std. Gas	0.079	12:18
07 Room Air	0.000	12:18

Lot No = 09216080A3
Cyl No = 25
Exp Date = 05/05/2018
County = 36 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = DEVL 8164.14.00 09/16
06/15/2017 12:26

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:27
02 Std. Gas	0.079	12:27
03 Room Air	0.000	12:28
04 Std. Gas	0.079	12:28
05 Room Air	0.000	12:29
06 Std. Gas	0.079	12:29
07 Room Air	0.000	12:30

Lot No = 09216080A3
Cyl No = 25
Exp Date = 05/05/2018
County = 36 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC