

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-002669

Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 115 psi Regulator: 125 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP3002 Lot #: 17160 Exp. Date: 6.8.19
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.400 AC
Sim. Ser #: MP3061 Lot #: 201605F Exp. Date: 5.19.18

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR 7089 Lot #: ICS 5 Exp. Date: 5.15.19
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 03417080A3 Cyl #: 27 Exp. Date: 3-5-2019

Test 1	<u>0.080</u>	Test 4	<u>0.080</u>	Test 7	<u>0.079</u>
Test 2	<u>0.080</u>	Test 5	<u>0.080</u>	Test 8	<u>0.079</u>
Test 3	<u>0.080</u>	Test 6	<u>0.080</u>	Test 9	<u>0.080</u>
Average	<u>0.080</u>				

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Eda
Breath Analyst Signature

4.4.2018
Date

Roberta Plueger-Nummers
Reviewed by

5 Apr 18
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = TOXL 8164.14.00 09/16
04/04/2018 14:49

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 03417080A3
Standard Cyl #? 27
Standard Expiration? 03/05/2019
Oper No? 666666

Flow Cal. Date: 09/09/2011
Slope 662
Intercept -510765

IR Calibration Date: 02/06/2015
 3um 9um

0th Coef(*100): -31426 -24477
1st Coef(*100): 275748 139415
2nd Coef(*100): 1078 912
H2O adj(mg/l*10k): 429 271

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = TOXL 8164.14.00 09/16
04/04/2018 14:57

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:58
02 Std. Sol.	0.017	14:59
03 Room Air	0.000	14:59
04 Std. Sol.	0.017	15:00
05 Room Air	0.000	15:00
06 Std. Sol.	0.017	15:01
07 Room Air	0.000	15:02

08 Sim Temp = 34.0°C

Simul Ser No = MP3002
Std Sol No = 17160
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = TOXL 8164.14.00 09/16
04/04/2018 15:08

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:08
02 Std. Sol.	0.391	15:09
03 Room Air	0.000	15:10
04 Std. Sol.	0.392	15:10
05 Room Air	0.000	15:11
06 Std. Sol.	0.392	15:12
07 Room Air	0.000	15:12

08 Sim Temp = 34.0°C

Simul Ser No = MP3061
Std Sol No = 201605F
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.400 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = TOXL 8164.14.00 09/16
04/04/2018 15:14

Test	AC	Time
01 Room Air	0.000	15:15
02 *Subject Test	INT*	15:15
03 Room Air	0.000	15:16

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 03417080A3
Cyl No = 27
Expiration Date = 03/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = TOXL 8164.14.00 09/16
04/04/2018 15:16

Test	AC	Time
01 Diagnostic	OK	15:17
02 Room Air	0.000	15:17
03 *Subject Test	RFI*	15:17
04 Room Air	0.000	15:18

*Invalid Test
 Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 03417080A3
Cyl No = 27
Expiration Date = 03/05/2019
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = TOXL 8164.14.00 09/16
04/04/2018 15:19

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:19
02 Std. Gas	0.080	15:20
03 Room Air	0.000	15:20
04 Std. Gas	0.080	15:20
05 Room Air	0.000	15:21
06 Std. Gas	0.080	15:21
07 Room Air	0.000	15:22

Lot No = 03417080A3
Cyl No = 27
Exp Date = 03/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = TOXL 8164.14.00 09/16
04/04/2018 15:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:23
02 Std. Gas	0.080	15:23
03 Room Air	0.000	15:23
04 Std. Gas	0.080	15:24
05 Room Air	0.000	15:24
06 Std. Gas	0.080	15:25
07 Room Air	0.000	15:25

Lot No = 03417080A3
Cyl No = 27
Exp Date = 03/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *CALIBRATION CHECK*
 0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = TOXL 8164.14.00 09/16
04/04/2018 15:25

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:26
02 Std. Gas	0.079	15:26
03 Room Air	0.000	15:27
04 Std. Gas	0.079	15:27
05 Room Air	0.000	15:28
06 Std. Gas	0.080	15:28
07 Room Air	0.000	15:28

Lot No = 03417080A3
Cyl No = 27
Exp Date = 03/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC