

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-003065

Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 52 psi Regulator: 50 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP3002 Lot #: 17160 Exp. Date: 6.8.2019
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.400 AC
Sim. Ser #: MP3061 Lot #: 201605F Exp. Date: 5.19.18

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR7089 Lot #: ICS 5 Exp. Date: 5.15.19
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 03417080A3 Cyl #: 27 Exp. Date: 3.5.19

Test 1	<u>0.080</u>	Test 4	<u>0.079</u>	Test 7	<u>0.079</u>
Test 2	<u>0.079</u>	Test 5	<u>0.079</u>	Test 8	<u>0.079</u>
Test 3	<u>0.079</u>	Test 6	<u>0.080</u>	Test 9	<u>0.080</u>
Average	<u>0.079</u>				

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Eder
Breath Analyst Signature

06 APR. 2018
Date

Roberta Iniguez-Nimmo
Reviewed by

06 APR. 2018
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003065
Location = TOXL 8164.14.00 09/16
04/06/2018 10:14

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:15
02 Std. Sol.	0.019	10:16
03 Room Air	0.000	10:16
04 Std. Sol.	0.019	10:17
05 Room Air	0.000	10:18
06 Std. Sol.	0.019	10:18
07 Room Air	0.000	10:19

08 Sim Temp = 34.0°C

Simul Ser No = MP3002

Std Sol No = 17160

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003065
Location = TOXL 8164.14.00 09/16
04/06/2018 10:22

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:22
02 Std. Sol.	0.390	10:23
03 Room Air	0.000	10:24
04 Std. Sol.	0.392	10:24
05 Room Air	0.000	10:25
06 Std. Sol.	0.391	10:26
07 Room Air	0.000	10:26

08 Sim Temp = 34.0°C

Simul Ser No = MP3061
Std Sol No = 201605F
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.400 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003065
Location = TOXL 8164.14.00 09/16
04/06/2018 10:29

Test	AC	Time
01 Room Air	0.000	10:30
02 *Subject Test	INT*	10:30
03 Room Air	0.000	10:31

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 03417080A3
Cyl No = 27
Expiration Date = 03/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10 AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003065
Location = TOXL 8164.14.00 09/16
04/06/2018 10:32

Test	AC	Time
01 Diagnostic	OK	10:33
02 Room Air	0.000	10:34
03 *Subject Test	RFI*	10:34
04 Room Air	0.000	10:34

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 03417080A3
Cyl No = 27
Expiration Date = 03/05/2019
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003065
Location = TOXL 8164.14.00 09/16
04/06/2018 10:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:35
02 Std. Gas	0.080	10:36
03 Room Air	0.000	10:36
04 Std. Gas	0.079	10:36
05 Room Air	0.000	10:37
06 Std. Gas	0.079	10:37
07 Room Air	0.000	10:38

Lot No = 03417080A3
Cyl No = 27
Exp Date = 03/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003065
Location = TOXL 8164.14.00 09/16
04/06/2018 10:38

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:39
02 Std. Gas	0.079	10:39
03 Room Air	0.000	10:40
04 Std. Gas	0.079	10:40
05 Room Air	0.000	10:41
06 Std. Gas	0.080	10:41
07 Room Air	0.000	10:42

Lot No = 03417080A3
Cyl No = 27
Exp Date = 03/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003065
Location = TOXL 8164.14.00 09/16
04/06/2018 10:45

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:46
02 Std. Gas	0.079	10:46
03 Room Air	0.000	10:47
04 Std. Gas	0.079	10:47
05 Room Air	0.000	10:48
06 Std. Gas	0.080	10:48
07 Room Air	0.000	10:49

Lot No = 03417080A3
Cyl No = 27
Exp Date = 03/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC