



ETHANOL GAS STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>Conner J Monk</i>		
Location <i>ASHL</i>	Intoxilyzer® Serial Number <i>80-005938</i>	
Gas Lot Number <i>34917080 A3</i>	Gas Cylinder Number <i>33</i>	Gas Expiration Date <i>2/5/2020</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.081* AC
 - B. *0.081* AC
 - C. *0.081* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Conner J Monk</i>	Date <i>4/16/2018</i>
Reviewed By (Crime Laboratory Use Only) <i>Robert J. Hume</i>	Date <i>4/9/2018</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005938
Location = ASHL 8164.14.00 09/16
04/06/2018 13:48

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:48
02 Std. Gas	0.081	13:49
03 Room Air	0.000	13:49
04 Std. Gas	0.081	13:50
05 Room Air	0.000	13:50
06 Std. Gas	0.081	13:50
07 Room Air	0.000	13:51

Lot No = 34917060A3
Cyl No = 33
Exp Date = 02/05/2020
County = 26 Oper No. = 131150



Operator Signature
CONNER MONK

Remarks: