

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-005940

Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 994 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.026 AC
Sim. Ser #: MP3002 Lot #: 17160 Exp. Date: 6-8-19
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.400 AC
Sim. Ser #: MP3061 Lot #: 201605F Exp. Date: 5-19-18

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR7089 Lot #: ICS#5 Exp. Date: 5-15-19
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 19817080A4 Cyl #: 21 Exp. Date: 9/5/19

Test 1	<u>0.080</u>	Test 4	<u>0.081</u>	Test 7	<u>0.081</u>
Test 2	<u>0.081</u>	Test 5	<u>0.080</u>	Test 8	<u>0.080</u>
Test 3	<u>0.080</u>	Test 6	<u>0.080</u>	Test 9	<u>0.080</u>
Average <u>0.080</u>					

C. Remarks/Maintenance: ANNUAL INSPECTION, EXTERNAL BATTERY
REPLACED AND TRAINED.

Instrument is acceptable to be used in the field.

Charles E. Ede
Breath Analyst Signature

4.9.2018
Date

Roberta Huque-Nimmo
Reviewed by

9 Apr 18
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.14.00 09/16
04/09/2018 14:41

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#%&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#%&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 19817080A4
Standard Cyl #? 21
Standard Expiration? 09/05/2019
Oper No? 666666

Flow Cal. Date: 06/21/2013
Slope 688
Intercept -659053

IR Calibration Date: 04/02/2015

	3um	9um
0th Coef(*100):	-18917	-25088
1st Coef(*100):	267616	136158
2nd Coef(*100):	3374	1493
H2O adj(mg/l*10k):	427	417

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks: PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.14.00 09/16
04/09/2018 14:45

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:45
02 Std. Sol.	0.018	14:46
03 Room Air	0.000	14:46
04 Std. Sol.	0.017	14:47
05 Room Air	0.000	14:48
06 Std. Sol.	0.018	14:48
07 Room Air	0.000	14:49

08 Sim Temp = 34.0°C

Simul Ser No = MP3002

Std Sol No = 17160

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.14.00 09/16
04/09/2018 14:50

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:51
02 Std. Sol.	0.398	14:52
03 Room Air	0.000	14:52
04 Std. Sol.	0.399	14:53
05 Room Air	0.000	14:54
06 Std. Sol.	0.398	14:54
07 Room Air	0.000	14:55

08 Sim Temp = 34.0°C

Simul Ser No = MP3061
Std Sol No = 201605F
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.400 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.14.00 09/16
04/09/2018 14:56

Test	AC	Time
01 Room Air	0.000	14:57
02 *Subject Test	INT*	14:57
03 Room Air	0.000	14:58

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 19817080A4
Cyl No = 21
Expiration Date = 09/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.14.00 09/16
04/09/2018 14:58

Test	AC	Time
01 Diagnostic	OK	14:59
02 Room Air	0.000	15:00
03 *Subject Test	RFI*	15:00
04 Room Air	0.000	15:00

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 19817080A4
Cyl No = 21
Expiration Date = 09/05/2019
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.14.00 09/16
04/09/2018 15:00

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:01
02 Std. Gas	0.080	15:01
03 Room Air	0.000	15:02
04 Std. Gas	0.081	15:02
05 Room Air	0.000	15:03
06 Std. Gas	0.080	15:03
07 Room Air	0.000	15:04

Lot No = 19817080A4
Cyl No = 21
Exp Date = 09/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.14.00 09/16
04/09/2018 15:04

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:04
02 Std. Gas	0.081	15:05
03 Room Air	0.000	15:05
04 Std. Gas	0.080	15:05
05 Room Air	0.000	15:06
06 Std. Gas	0.080	15:06
07 Room Air	0.000	15:07

Lot No = 19817080A4
Cyl No = 21
Exp Date = 09/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = TOXL 8164.14.00 09/16
04/09/2018 15:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:08
02 Std. Gas	0.081	15:08
03 Room Air	0.000	15:08
04 Std. Gas	0.080	15:09
05 Room Air	0.000	15:09
06 Std. Gas	0.080	15:10
07 Room Air	0.000	15:10

Lot No = 19817080A4
Cyl No = 21
Exp Date = 09/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC