

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-005951

Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 22 psi Regulator: 25 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP3002 Lot #: 17160 Exp. Date: 6-8-19
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.400 AC
Sim. Ser #: MP3061 Lot #: 201605F Exp. Date: 5-19-18

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR7089 Lot #: ICS 5 Exp. Date: 5-15-19
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 19817080A4 Cyl #: 21 Exp. Date: 9/5/2019

Test 1 <u>0.083</u>	Test 4 <u>0.083</u>	Test 7 <u>0.082</u>
Test 2 <u>0.083</u>	Test 5 <u>0.083</u>	Test 8 <u>0.082</u>
Test 3 <u>0.083</u>	Test 6 <u>0.082</u>	Test 9 <u>0.082</u>
Average <u>0.083</u>		

C. Remarks/Maintenance: ANNUAL INSPECTION, REPLACED BREATH HOSE O-RING.

Instrument is acceptable to be used in the field.

Charles E. Eda
Breath Analyst Signature

4.6.2018
Date

Roberta Mungu-nimmo
Reviewed by

9 Apr 18
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005951
Location = TOXL 8164.14.00 09/16
04/06/2018 15:03

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#%&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#%&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer (Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 03417080A3
Standard Cyl #? 27
Standard Expiration? 03/05/2019
Oper No? 666666

Flow Cal. Date: 06/21/2013
Slope 693
Intercept -721277

IR Calibration Date: 05/08/2017

	3um	9um
0th Coef (*100):	-17201	-17496
1st Coef (*100):	267144	137813
2nd Coef (*100):	2948	1389
H2O adj (mg/l*10k):	438	428

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005951
Location = TOXL 8164.14.00 09/16
04/06/2018 15:07

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:08
02 Std. Sol.	0.020	15:08
03 Room Air	0.000	15:09
04 Std. Sol.	0.021	15:10
05 Room Air	0.000	15:10
06 Std. Sol.	0.021	15:11
07 Room Air	0.000	15:11

08 Sim Temp = 34.0°C

Simul Ser No = MP3002
Std Sol No = 17160
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005951
Location = TOXL 8164.14.00 09/16
04/06/2018 15:12

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:13
02 Std. Sol.	0.404	15:13
03 Room Air	0.000	15:14
04 Std. Sol.	0.407	15:15
05 Room Air	0.000	15:15
06 Std. Sol.	0.405	15:16
07 Room Air	0.000	15:17

08 Sim Temp = 34.0°C

Simul Ser No = MP3061
Std Sol No = 201605F
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: HIGH AC
0.400 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005951
Location = TOXL 8164.14.00 09/16
04/06/2018 15:18

Test	AC	Time
01 Room Air	0.000	15:19
02 *Subject Test	INT*	15:20
03 Room Air	0.000	15:20

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 03417080A3
Cyl No = 27
Expiration Date = 03/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005951
Location = TOXL 8164.14.00 09/16
04/06/2018 15:21

Test	AC	Time
01 Diagnostic	OK	15:22
02 Room Air	0.000	15:23
03 *Subject Test	RFI*	15:23
04 Room Air	0.000	15:23

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 19817080A4
Cyl No = 21
Expiration Date = 09/05/2019
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005951
Location = TOXL 8164.14.00 09/16
04/06/2018 15:24

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:25
02 Std. Gas	0.083	15:25
03 Room Air	0.000	15:26
04 Std. Gas	0.083	15:26
05 Room Air	0.000	15:27
06 Std. Gas	0.083	15:27
07 Room Air	0.000	15:28

Lot No = 19817080A4
Cyl No = 21
Exp Date = 09/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005951
Location = TOXL 8164.14.00 09/16
04/06/2018 15:28

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:28
02 Std. Gas	0.083	15:29
03 Room Air	0.000	15:29
04 Std. Gas	0.083	15:30
05 Room Air	0.000	15:30
06 Std. Gas	0.082	15:30
07 Room Air	0.000	15:31

Lot No = 19817080A4
Cyl No = 21
Exp Date = 09/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005951
Location = TOXL 8164.14.00 09/16
04/06/2018 15:31

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:32
02 Std. Gas	0.082	15:32
03 Room Air	0.000	15:33
04 Std. Gas	0.082	15:33
05 Room Air	0.000	15:33
06 Std. Gas	0.082	15:34
07 Room Air	0.000	15:34

Lot No = 19817080A4
Cyl No = 21
Exp Date = 09/05/2019
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000