

**INTOXILYZER® 8000 INSPECTION**

Intoxilyzer® 8000 Serial Number: 80-004204

Location: TOL

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 344 psi Regulator: 350 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: MP3060 Lot #: 16200 Exp. Date: 17 Aug 18
- b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: MP3062 Lot #: 17350 Exp. Date: 11 Oct 19

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR5144 Lot #: 1CS5 Exp. Date: 15 May 19  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value \_\_\_\_\_ AC  
Gas Cyl Lot #: 20016080A2 Cyl #: 48 Exp. Date: 5 Sep 18

Test 1 <u>0.080</u>	Test 4 <u>0.080</u>	Test 7 <u>0.081</u>
Test 2 <u>0.080</u>	Test 5 <u>0.081</u>	Test 8 <u>0.080</u>
Test 3 <u>0.080</u>	Test 6 <u>0.080</u>	Test 9 <u>0.080</u>
Average <u>0.080</u>		

C. Remarks/Maintenance: Annual inspection. Replaced o-rings on breath hose and simulator return.

Instrument is acceptable to be used in the field.

Roberta Munoz-Romero  
Breath Analyst Signature

10 Apr 18  
Date

Charles E. Edr  
Reviewed by

10 APR 2018  
Date



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004204  
Location = TOXL      8164.14.00 09/16  
04/10/2018      10:36

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:37
02 Std. Sol.	0.019	10:38
03 Room Air	0.000	10:38
04 Std. Sol.	0.020	10:39
05 Room Air	0.000	10:39
06 Std. Sol.	0.020	10:40
07 Room Air	0.000	10:41

08 Sim Temp = 34.0°C

Simul Ser No = MP3060  
Std Sol No = 16200  
County = 08      Oper No. = 888888



Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: low AC  
0.020AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004204  
Location = TOXL      8164.14.00 09/16  
04/10/2018      10:54

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:55
02 Std. Sol.	0.294	10:56
03 Room Air	0.000	10:56
04 Std. Sol.	0.295	10:57
05 Room Air	0.000	10:57
06 Std. Sol.	0.295	10:58
07 Room Air	0.000	10:59


08 Sim Temp = 34.0°C

Simul Ser No = MP3062

Std Sol No = 17350

County = 08

Oper No. = 888888



Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: High AC  
0.300AC

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004204  
Location = TOXL                8164.14.00 09/16  
04/10/2018                        10:31

Test	AC	Time
01 Room Air	0.000	10:32
02 *Subject Test	INT*	10:32
03 Room Air	0.000	10:33

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = INTERFERENT  
Dr. Lic. = ND/DIS821456  
Lot No = 20016080A2  
Cyl No = 48  
Expiration Date = 09/05/2018  
County = 08                            Oper No. = 888888



Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: *Interferent Check*  
*0.10 AC + 0.05% Acetone*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004204  
Location = TOXL                    8164.14.00 09/16  
04/10/2018                            10:33

Test	AC	Time
01 Diagnostic	OK	10:34
02 Room Air	0.000	10:35
03 *Subject Test	RFI*	10:35
04 Room Air	0.000	10:35

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI CHECK  
Dr. Lic. = ND/DIS821456  
Lot No = 20016080A2  
Cyl No = 48  
Expiration Date = 09/05/2018  
County = 08                            Oper No. = 888888

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: *RFI Check*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004204  
Location = TOXL      8164.14.00 09/16  
04/10/2018      10:13

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:14
02 Std. Gas	0.080	10:14
03 Room Air	0.000	10:15
04 Std. Gas	0.080	10:15
05 Room Air	0.000	10:16
06 Std. Gas	0.080	10:16
07 Room Air	0.000	10:17

Lot No = 20016080A2  
Cyl No = 48  
Exp Date = 09/05/2018  
County = 08      Oper No. = 888888



Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: *Calibration Check*

*0.080 AC*

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004204  
Location = TOXL      8164.14.00 09/16  
04/10/2018      10:24

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:24
02 Std. Gas	0.080	10:25
03 Room Air	0.000	10:25
04 Std. Gas	0.081	10:26
05 Room Air	0.000	10:26
06 Std. Gas	0.080	10:27
07 Room Air	0.000	10:27

Lot No = 20016080A2  
Cyl No = 48  
Exp Date = 09/05/2018  
County = 08      Oper No. = 888888

  
Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: Calibration Check  
0.080 AC  
Form 106-I8000




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004204  
Location = TOXL                    8164.14.00 09/16  
04/10/2018                            10:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:28
02 Std. Gas	0.081	10:28
03 Room Air	0.000	10:29
04 Std. Gas	0.080	10:29
05 Room Air	0.000	10:30
06 Std. Gas	0.080	10:30
07 Room Air	0.000	10:31

Lot No = 20016080A2  
Cyl No = 48  
Exp Date = 09/05/2018  
County = 08                            Oper No. = 888888

  
Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: Calibration Check  
                  0.080 AC  
                  Form 106-I8000